


LB 131857

SFAM01.1

INSTRUMENT ID: cv1

μ Abs.:

42285



Linear

A= 0.0000e+000

B= 2.3335e-004 *slope*

C= -4.6694e-002 *y-intercept*

Rho= 0.9993717

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>0/01</i>
0.00	0.000	-0.024	-0.024	99	0.000	99					
0.05	0.050					442					
0.20	0.200	0.200	-0.000	1057	0.0 %	1057	0				
2.50	2.500	2.489	-0.011	10868	0.0 %	10868					
5.00	5.000	4.954	-0.046	21429	0.0 %	21429					
7.50	7.500	7.760	0.260	33455	0.0 %	33455					
10.0	10.000	9.821	-0.179	42285	0.0 %	42285					

0/01
= 0
- 1
3
- 2

LB131857

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	99	-		0 SFAM01.1	PPB	8/2/2024 17:06	Std
	0.2 S01	1057	-		0.2 SFAM01.1	PPB	8/2/2024 17:08	Std
	2.5 S02	10868	-		2.5 SFAM01.1	PPB	8/2/2024 17:10	Std
	5 S03	21429	-		5 SFAM01.1	PPB	8/2/2024 17:13	Std
	7.5 S04	33455	-		7.5 SFAM01.1	PPB	8/2/2024 17:15	Std
	10 S05	42285	-		10 SFAM01.1	PPB	8/2/2024 17:17	Std
ICV058	ICV058	16548	3.8148 -		SFAM01.1	PPB	8/2/2024 17:23	SMPL
ICB058	ICB058	-15	-0.0502 -		SFAM01.1	PPB	8/2/2024 17:25	SMPL
CCV086	CCV086	21895	5.0625 -		SFAM01.1	PPB	8/2/2024 17:27	SMPL
CCB086	CCB086	-48	-0.0579 -		SFAM01.1	PPB	8/2/2024 17:30	SMPL
PB162480BL	PBS480	86	-0.0266 -		SFAM01.1	PPB	8/2/2024 17:32	SMPL
P3273-01	MDCVA6	1253	0.2457 -		SFAM01.1	PPB	8/2/2024 17:34	SMPL
P3273-02	MDCVB4	698	0.1162 -		SFAM01.1	PPB	8/2/2024 17:37	SMPL
P3273-03	MDCVC6	732	0.1241 -		SFAM01.1	PPB	8/2/2024 17:39	SMPL
P3273-04	MDCVC7	584	0.0896 -		SFAM01.1	PPB	8/2/2024 17:41	SMPL
P3273-05	MDCVC8	627	0.0996 -		SFAM01.1	PPB	8/2/2024 17:43	SMPL
P3273-06	MDCVC9	591	0.0912 -		SFAM01.1	PPB	8/2/2024 17:46	SMPL
P3273-07	MDCVD0	681	0.1122 -		SFAM01.1	PPB	8/2/2024 17:48	SMPL
P3273-08	MDCVD1	701	0.1169 -		SFAM01.1	PPB	8/2/2024 17:50	SMPL
P3273-09	MDCVD2	762	0.1311 -		SFAM01.1	PPB	8/2/2024 17:52	SMPL
P3273-10	MDCVD3	600	0.0933 -		SFAM01.1	PPB	8/2/2024 17:55	SMPL
P3273-11	MDCVD4	659	0.1071 -		SFAM01.1	PPB	8/2/2024 17:57	SMPL
P3273-12	MDCVE1	235	0.0081 -		SFAM01.1	PPB	8/2/2024 17:59	SMPL
P3273-13	MDCVE2	722	0.1218 -		SFAM01.1	PPB	8/2/2024 18:01	SMPL
P3273-14	MDCVE3	827	0.1463 -		SFAM01.1	PPB	8/2/2024 18:04	SMPL
P3273-15	MDCVE7	771	0.1332 -		SFAM01.1	PPB	8/2/2024 18:06	SMPL
P3273-16	MDCVE8	696	0.1157 -		SFAM01.1	PPB	8/2/2024 18:08	SMPL
P3273-17	MDCVE9	786	0.1367 -		SFAM01.1	PPB	8/2/2024 18:10	SMPL
P3273-18	MDCVF7	864	0.1549 -		SFAM01.1	PPB	8/2/2024 18:13	SMPL
P3273-19	MDCVK4	766	0.1321 -		SFAM01.1	PPB	8/2/2024 18:15	SMPL
CCV087	CCV087	22142	5.1201 -		SFAM01.1	PPB	8/2/2024 18:17	SMPL
CCB087	CCB087	-49	-0.0581 -		SFAM01.1	PPB	8/2/2024 18:20	SMPL
P3273-20	MDCVK4D	836	0.1484 -		SFAM01.1	PPB	8/2/2024 18:22	SMPL
P3273-21	MDCVK4S	11769	2.6996 -		SFAM01.1	PPB	8/2/2024 18:24	SMPL
PB162482BL	PBS482	-18	-0.0509 -		SFAM01.1	PPB	8/2/2024 18:26	SMPL
P3274-01	MDCVK5	689	0.1141 -		SFAM01.1	PPB	8/2/2024 18:29	SMPL
P3274-02	MDCVK6	285	0.0198 -		SFAM01.1	PPB	8/2/2024 18:31	SMPL
P3274-03	MDCVK7	414	0.0499 -		SFAM01.1	PPB	8/2/2024 18:33	SMPL
P3274-04	MDCVK9	707	0.1183 -		SFAM01.1	PPB	8/2/2024 18:35	SMPL
P3274-05	MDCVA5	529	0.0767 -		SFAM01.1	PPB	8/2/2024 18:38	SMPL
P3274-06	MDCVA7	325	0.0291 -		SFAM01.1	PPB	8/2/2024 18:40	SMPL
P3274-07	MDCVB7	595	0.0921 -		SFAM01.1	PPB	8/2/2024 18:42	SMPL
P3274-08	MDCVB8	715	0.1202 -		SFAM01.1	PPB	8/2/2024 18:45	SMPL
P3274-09	MDCVB9	528	0.0765 -		SFAM01.1	PPB	8/2/2024 18:47	SMPL
P3274-10	MDCVC5	729	0.1234 -		SFAM01.1	PPB	8/2/2024 18:49	SMPL
P3274-11	MDCVF6	392	0.0448 -		SFAM01.1	PPB	8/2/2024 18:51	SMPL

LB131857

INSTRUMENT ID : CV1

P3274-12	MDCVF8	517	0.0739 -	SFAM01.1	PPB	8/2/2024 18:54 SMPL
P3274-13	MDCVF8D	458	0.0602 -	SFAM01.1	PPB	8/2/2024 18:56 SMPL
P3274-14	MDCVF8S	11235	2.575 -	SFAM01.1	PPB	8/2/2024 18:58 SMPL
P3275-01	MDCVD5	577	0.0879 -	SFAM01.1	PPB	8/2/2024 19:00 SMPL
P3275-02	MDCVD6	752	0.1288 -	SFAM01.1	PPB	8/2/2024 19:03 SMPL
P3275-03	MDCVD7	554	0.0826 -	SFAM01.1	PPB	8/2/2024 19:05 SMPL
CCV088	CCV088	21898	5.0632 -	SFAM01.1	PPB	8/2/2024 19:07 SMPL
CCB088	CCB088	-67	-0.0623 -	SFAM01.1	PPB	8/2/2024 19:09 SMPL
P3275-04	MDCVD8	1142	0.2198 -	SFAM01.1	PPB	8/2/2024 19:12 SMPL
P3275-05	MDCVD9	1109	0.2121 -	SFAM01.1	PPB	8/2/2024 19:14 SMPL
P3275-06	MDCVE0	963	0.178 -	SFAM01.1	PPB	8/2/2024 19:16 SMPL
P3275-07	MDCVE0D	778	0.1349 -	SFAM01.1	PPB	8/2/2024 19:19 SMPL
P3275-08	MDCVE0S	12752	2.929 -	SFAM01.1	PPB	8/2/2024 19:21 SMPL
CCV089	CCV089	22094	5.1089 -	SFAM01.1	PPB	8/2/2024 19:23 SMPL
CCB089	CCB089	-78	-0.0649 -	SFAM01.1	PPB	8/2/2024 19:25 SMPL