

# Analytical Summary Report

Analysis Method: 7196A

ANALYST: rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: jignesh

Run Number: LB131861

pH Meter ID: WC pH Meter-1

| Reagent/Standard                  | Lot/Log # |
|-----------------------------------|-----------|
| hexavalent chromium color reagent | WP108907  |
| 5N sulfuric acid                  | WP107791  |
| HNO3 Hex-Chrome, 5M               | WP107796  |
| Hexchrome Cleaning Solution       | WP108645  |

Intercept: 0.0012

Slope: 0.7659

Regression: 0.999993

| Seq | Lab ID | True Value (mg/l) | DF | Initial Vol (ml) | Final Vol (ml) | pH HNO3 | pH H2SO4 | Absorb.at 540nm |       | Absorbance Difference | Result (mg/L) | %D   | Anal Date  | Anal Time |
|-----|--------|-------------------|----|------------------|----------------|---------|----------|-----------------|-------|-----------------------|---------------|------|------------|-----------|
|     |        |                   |    |                  |                |         |          | Backgrnd        | Color |                       |               |      |            |           |
| 1   | CAL1   | 0                 | 1  | 100              | 100            | 7.28    | 1.74     | 0.000           | 0.000 | 0.000                 | -0.00         |      | 08/05/2024 | 12:30     |
| 2   | CAL2   | 0.01              | 1  | 100              | 100            | 7.36    | 1.89     | 0.000           | 0.009 | 0.009                 | 0.010         | 0    | 08/05/2024 | 12:31     |
| 3   | CAL3   | 0.025             | 1  | 100              | 100            | 7.38    | 1.86     | 0.000           | 0.020 | 0.020                 | 0.024         | -4   | 08/05/2024 | 12:32     |
| 4   | CAL4   | 0.05              | 1  | 100              | 100            | 7.38    | 1.88     | 0.000           | 0.039 | 0.039                 | 0.049         | -2   | 08/05/2024 | 12:33     |
| 5   | CAL5   | 0.1               | 1  | 100              | 100            | 7.35    | 1.86     | 0.000           | 0.080 | 0.080                 | 0.102         | 2    | 08/05/2024 | 12:34     |
| 6   | CAL6   | 0.5               | 1  | 100              | 100            | 7.39    | 1.91     | 0.000           | 0.384 | 0.384                 | 0.499         | -0.2 | 08/05/2024 | 12:35     |
| 7   | CAL7   | 1                 | 1  | 100              | 100            | 7.36    | 1.88     | 0.000           | 0.767 | 0.767                 | 0.999         | -0.1 | 08/05/2024 | 12:36     |



## Analytical Summary Report

Analysis Method: 7196A

ANALYST: rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: jignesh

Run Number: LB131861

pH Meter ID: WC pH Meter-1

| Seq | Lab ID     | True Value | DF | Initial Vol (ml/gm) | Final Vol (ml) | pH HN03 | pH H2SO4 | Absorb.at540nm |       | Absorbance Difference | Intermediate Result (mg/L) | Anal Date  | Anal Time |
|-----|------------|------------|----|---------------------|----------------|---------|----------|----------------|-------|-----------------------|----------------------------|------------|-----------|
|     |            |            |    |                     |                |         |          | Backgrnd       | Color |                       |                            |            |           |
| 1   | ICV        | 0.5        | 1  | 100                 | 100            | 7.43    | 1.93     | 0.000          | 0.386 | 0.386                 | 0.502                      | 08/05/2024 | 12:37     |
| 2   | ICB        |            | 1  | 100                 | 100            | 7.25    | 1.77     | 0.000          | 0.000 | 0.000                 | -0.002                     | 08/05/2024 | 12:38     |
| 3   | CCV1       | 0.5        | 1  | 100                 | 100            | 7.45    | 1.96     | 0.000          | 0.383 | 0.383                 | 0.498                      | 08/05/2024 | 12:39     |
| 4   | CCB1       |            | 1  | 100                 | 100            | 7.27    | 1.74     | 0.000          | 0.001 | 0.001                 | 0.000                      | 08/05/2024 | 12:40     |
| 5   | RL Check   | 0.01       | 1  | 100                 | 100            | 7.40    | 1.91     | 0.000          | 0.009 | 0.009                 | 0.010                      | 08/05/2024 | 12:41     |
| 6   | PB162492BL |            | 1  | 2.50                | 100            | 7.30    | 1.79     | 0.000          | 0.000 | 0.000                 | -0.002                     | 08/05/2024 | 12:42     |
| 7   | PB162492BS | 20         | 1  | 2.50                | 100            | 7.42    | 1.94     | 0.000          | 0.384 | 0.384                 | 0.500                      | 08/05/2024 | 12:43     |
| 8   | P3448-01   |            | 1  | 2.53                | 100            | 7.60    | 2.09     | 0.010          | 0.012 | 0.002                 | 0.001                      | 08/05/2024 | 12:44     |
| 9   | P3448-03   |            | 1  | 2.54                | 100            | 7.68    | 2.16     | 0.008          | 0.008 | 0.000                 | -0.002                     | 08/05/2024 | 12:45     |
| 10  | P3448-05   |            | 1  | 2.52                | 100            | 7.60    | 2.20     | 0.014          | 0.014 | 0.000                 | -0.002                     | 08/05/2024 | 12:46     |
| 11  | P3448-07   |            | 1  | 2.51                | 100            | 7.54    | 2.04     | 0.010          | 0.010 | 0.000                 | -0.002                     | 08/05/2024 | 12:47     |
| 12  | P3448-09   |            | 1  | 2.54                | 100            | 7.60    | 2.12     | 0.008          | 0.009 | 0.001                 | 0.000                      | 08/05/2024 | 12:48     |
| 13  | P3448-11   |            | 1  | 2.53                | 100            | 7.58    | 2.18     | 0.014          | 0.014 | 0.000                 | -0.002                     | 08/05/2024 | 12:49     |
| 14  | P3448-13   |            | 1  | 2.51                | 100            | 7.60    | 2.06     | 0.012          | 0.012 | 0.000                 | -0.002                     | 08/05/2024 | 12:50     |
| 15  | P3463-01   |            | 1  | 2.54                | 100            | 7.52    | 2.10     | 0.006          | 0.007 | 0.001                 | 0.000                      | 08/05/2024 | 12:51     |
| 16  | CCV2       | 0.5        | 1  | 100                 | 100            | 7.42    | 1.93     | 0.000          | 0.384 | 0.384                 | 0.500                      | 08/05/2024 | 12:52     |
| 17  | CCB2       |            | 1  | 100                 | 100            | 7.29    | 1.75     | 0.000          | 0.000 | 0.000                 | -0.002                     | 08/05/2024 | 12:53     |
| 18  | P3463-03   |            | 1  | 2.51                | 100            | 7.54    | 2.04     | 0.006          | 0.006 | 0.000                 | -0.002                     | 08/05/2024 | 12:54     |
| 19  | P3463-03DU |            | 1  | 2.52                | 100            | 7.52    | 2.08     | 0.005          | 0.006 | 0.001                 | 0.000                      | 08/05/2024 | 12:55     |
| 20  | P3463-03MS | 40         | 2  | 2.51                | 100            | 7.58    | 2.10     | 0.001          | 0.314 | 0.313                 | 0.407                      | 08/05/2024 | 12:56     |
| 21  | P3463-03MS | 1284       | 40 | 2.52                | 100            | 7.52    | 2.06     | 0.000          | 0.610 | 0.610                 | 0.795                      | 08/05/2024 | 12:57     |
| 22  | P3463-03MS | 40         | 2  | 2.51                | 100            | 7.60    | 2.06     | 0.002          | 0.380 | 0.378                 | 0.492                      | 08/05/2024 | 12:58     |
| 23  | CCV3       | 0.5        | 1  | 100                 | 100            | 7.46    | 1.92     | 0.000          | 0.386 | 0.386                 | 0.502                      | 08/05/2024 | 12:59     |
| 24  | CCB3       |            | 1  | 100                 | 100            | 7.24    | 1.74     | 0.000          | 0.000 | 0.000                 | -0.002                     | 08/05/2024 | 13:00     |