

LS131874


SFAM01.1

INSTRUMENT ID: EV1

Linear

μ Abs.:

5065E



A= 0.0000e+000

B= 1.9752e-004 slope

C= -1.0354e-002 y-intercept

Rho= 0.9998821

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.012	0.012	114	0.000	114					
0.05	0.050					442	0				2
0.20	0.200	0.185	-0.015	989	0.0 %	989					2
2.50	2.500	2.448	-0.052	12445	0.0 %	12445					2
5.00	5.000	5.115	0.115	25947	0.0 %	25947					1
7.50	7.500	7.445	-0.055	37747	0.0 %	37747					0
10.0	10.000	9.995	-0.005	50656	0.0 %	50656					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	114	-		0 SFAM01.1	PPB	8/5/2024 17:47	Std
	0.2 S01	989	-		0.2 SFAM01.1	PPB	8/5/2024 17:49	Std
	2.5 S02	12445	-		2.5 SFAM01.1	PPB	8/5/2024 17:51	Std
	5 S03	25947	-		5 SFAM01.1	PPB	8/5/2024 17:53	Std
	7.5 S04	37747	-		7.5 SFAM01.1	PPB	8/5/2024 17:56	Std
	10 S05	50656	-		10 SFAM01.1	PPB	8/5/2024 17:58	Std
ICV061	ICV061	19787	3.8979 -		SFAM01.1	PPB	8/5/2024 18:01	SMPL
ICB061	ICB061	-3	-0.0109 -		SFAM01.1	PPB	8/5/2024 18:03	SMPL
CCV097	CCV097	25644	5.0548 -		SFAM01.1	PPB	8/5/2024 18:05	SMPL
CCB097	CCB097	-68	-0.0238 -		SFAM01.1	PPB	8/5/2024 18:08	SMPL
PB162515BL	PBS515	71	0.0037 -		SFAM01.1	PPB	8/5/2024 18:10	SMPL
P3269-01	MA4D96	319784	HIGH -		SFAM01.1	PPB	8/5/2024 18:12	
P3269-02	MA4D22	10535	2.0705 -		SFAM01.1	PPB	8/5/2024 18:14	SMPL
P3269-03	MA4D39	12017	2.3632 -		SFAM01.1	PPB	8/5/2024 18:16	SMPL
P3269-04	MA4D40	9242	1.8151 -		SFAM01.1	PPB	8/5/2024 18:19	SMPL
P3269-05	MA4D41	13639	2.6836 -		SFAM01.1	PPB	8/5/2024 18:21	SMPL
P3269-06	MA4D42	2967	0.5757 -		SFAM01.1	PPB	8/5/2024 18:23	SMPL
P3269-07	MA4D42D	3265	0.6345 -		SFAM01.1	PPB	8/5/2024 18:25	SMPL
P3269-08	MA4D42S	14830	2.9188 -		SFAM01.1	PPB	8/5/2024 18:28	SMPL
P3285-01	MA4CM2	1439	0.2739 -		SFAM01.1	PPB	8/5/2024 18:30	SMPL
P3285-02	MA4D97	300287	HIGH -		SFAM01.1	PPB	8/5/2024 18:32	
P3285-03	MA4CM2D	1433	0.2727 -		SFAM01.1	PPB	8/5/2024 18:40	SMPL
P3285-04	MA4CM2S	11974	2.3547 -		SFAM01.1	PPB	8/5/2024 18:42	SMPL
CCV098	CCV098	22825	4.498 -		SFAM01.1	PPB	8/5/2024 18:55	SMPL
CCB098	CCB098	-145	-0.039 -		SFAM01.1	PPB	8/5/2024 18:58	SMPL
P3269-01DLX25	MA4D96	20122	3.9641 -		SFAM01.1	PPB	8/5/2024 19:02	SMPL
P3285-02DLX25	MA4D97	19082	3.7587 -		SFAM01.1	PPB	8/5/2024 19:16	SMPL
CCV099	CCV099	22386	4.4113 -		SFAM01.1	PPB	8/5/2024 19:18	SMPL
CCB099	CCB099	-78	-0.0258 -		SFAM01.1	PPB	8/5/2024 19:21	SMPL