

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131877

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/06/2024	11:10
2	CAL2	1	Water	NA	NA	20.1	7.00	08/06/2024	11:11
3	CAL3	1	Water	NA	NA	20.3	10.02	08/06/2024	11:15
4	ICV	1	Water	NA	NA	20.4	7.01	08/06/2024	11:20
5	CCV1	1	Water	NA	NA	20.2	2.01	08/06/2024	11:25
6	P3474-01	1	Solid	20.02	20	22.7	5.47	08/06/2024	11:30
7	P3474-04	1	Solid	20.01	20	22.6	5.64	08/06/2024	11:35
8	P3474-07	1	Solid	20.03	20	22.9	5.78	08/06/2024	11:44
9	P3474-10	1	Solid	20.02	20	22.1	7.23	08/06/2024	11:50
10	P3474-10DUP	1	Solid	20.03	20	22.2	7.24	08/06/2024	11:51
11	CCV2	1	Water	NA	NA	20.3	12.02	08/06/2024	12:00

WORKLIST(Hardcopy Internal Chain)

131277

WorkList Name : ph p3474

WorkList ID : 182384

Department : Wet-Chemistry

Date : 08-06-2024 10:46:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3474-01	WC-A1-COMP	Solid	pH	Cool 4 deg C	EART12	D21	08/05/2024	9045D
P3474-04	WC-B1-COMP	Solid	pH	Cool 4 deg C	EART12	D21	08/05/2024	9045D
P3474-07	WC-AZ-COMP	Solid	pH	Cool 4 deg C	EART12	D21	08/05/2024	9045D
P3474-10	WC-B2-COMP	Solid	pH	Cool 4 deg C	EART12	D21	08/05/2024	9045D

Date/Time 08/06/24 11:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 08/06/24 13:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]