

LB131889

7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

46871

A= 0.0000e+000

B= 2.1744e-004

C= 3.3038e-002

Rho= 0.9989966

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	0.080	0.080	217	0.000	217					
0.2	0.200	0.207	0.007	802	0.0 %	802					
2.5	2.500	2.514	0.014	11410	0.0 %	11410					
5.0	5.000	5.000	-0.000	22842	0.0 %	22842					
7.5	7.500	7.173	-0.327	32838	0.0 %	32838					
10.0	10.000	10.225	0.225	46871	0.0 %	46871					

0/0 D
3
1
6
-4
2

LB131889 IONSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Aug 2024 10:47:08

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	217	-	0.0000	7470A	PPB	07 Aug 2024 10:51:27	S	Std
0.2 - 1	SO-2	802	-	0.2000	7470A	PPB	07 Aug 2024 10:53:44	S	Std
2.5 - 1	52.5	11410	-	2.5000	7470A	PPB	07 Aug 2024 10:56:01	S	Std
5.0 - 1	55	22842	-	5.0000	7470A	PPB	07 Aug 2024 10:58:18	S	Std
7.5 - 1	57.5	32838	-	7.5000	7470A	PPB	07 Aug 2024 11:00:35	S	Std
10.0 - 1	510	46871	-	10.0000	7470A	PPB	07 Aug 2024 11:05:29	S	Std
ICV63 - 1	ICV63	17062	3.7431	-	7470A	PPB	07 Aug 2024 11:08:15	U	SMPL
ICB63 - 1	ICB63	-210	-0.0126	-	7470A	PPB	07 Aug 2024 11:10:30	U	SMPL
CCV04 - 1	CCV04	21137	4.6292	-	7470A	PPB	07 Aug 2024 11:12:47	U	SMPL
CCB04 - 1	CCB04	-23	0.0280	-	7470A	PPB	07 Aug 2024 11:16:59	U	SMPL
CRA - 1	CRA	891	0.2268	-	7470A	PPB	07 Aug 2024 11:19:14	U	SMPL
HighStd - 1	HighStd	47051	10.2640	-	7470A	PPB	07 Aug 2024 11:21:29	U	SMPL
ChkStd - 1	ChkStd	29934	6.5420	-	7470A	PPB	07 Aug 2024 11:23:45	U	SMPL
PB162531BL - 1	PBW	-303	-0.0328	-	7470A	PPB	07 Aug 2024 11:26:03	U	SMPL
PB162531BS - 1	LCSW	16057	3.5245	-	7470A	PPB	07 Aug 2024 11:28:21	U	SMPL
P3341-03 - 1	AU-05-072324	-243	-0.0198	-	7470A	PPB	07 Aug 2024 11:30:37	U	SMPL
P3462-04 - 1	JC-1-4-5	438	0.1283	-	7470A	PPB	07 Aug 2024 11:32:55	U	SMPL
P3462-08 - 1	JC-2	665	0.1776	-	7470A	PPB	07 Aug 2024 11:35:12	U	SMPL
P3462-12 - 1	JC-3-6-7	113	0.0578	-	7470A	PPB	07 Aug 2024 11:37:29	U	SMPL
P3474-02 - 1	WC-A1-COMP	-87	0.0141	-	7470A	PPB	07 Aug 2024 11:39:46	U	SMPL
CCV05 - 1	CCV05	20982	4.5955	-	7470A	PPB	07 Aug 2024 11:42:01	U	SMPL
CCB05 - 1	CCB05	-170	-0.0039	-	7470A	PPB	07 Aug 2024 11:44:17	U	SMPL
P3474-05 - 1	WC-B1-COMP	40	0.0417	-	7470A	PPB	07 Aug 2024 11:46:34	U	SMPL
P3474-08 - 1	WC-A2-COMP	104	0.0557	-	7470A	PPB	07 Aug 2024 11:48:48	U	SMPL
P3474-11 - 1	WC-B2-COMP	156	0.0670	-	7470A	PPB	07 Aug 2024 11:51:03	U	SMPL
P3474-11DUP - 1	WC-B2-COMPDUP	130	0.0613	-	7470A	PPB	07 Aug 2024 11:53:19	U	SMPL
P3474-11MS - 1	WC-B2-COMPMS	16557	3.6333	-	7470A	PPB	07 Aug 2024 11:55:34	U	SMPL
P3474-11MSD - 1	WC-B2-COMPMSD	15342	3.3691	-	7470A	PPB	07 Aug 2024 11:57:51	U	SMPL
PB162557BL - 1	PBW	-50	0.0222	-	7470A	PPB	07 Aug 2024 12:00:09	U	SMPL
PB162557BS - 1	LCSW	14984	3.2912	-	7470A	PPB	07 Aug 2024 12:02:28	U	SMPL
P3455-01 - 1	RW7-SP100-50-20240801	7	0.0346	-	7470A	PPB	07 Aug 2024 12:04:44	U	SMPL
P3455-01DUP - 1	RW7-SP100-50-20240801DUP	198	0.0761	-	7470A	PPB	07 Aug 2024 12:07:01	U	SMPL
CCV06 - 1	CCV06	21456	4.6985	-	7470A	PPB	07 Aug 2024 12:09:17	U	SMPL
CCB06 - 1	CCB06	-258	-0.0231	-	7470A	PPB	07 Aug 2024 12:11:32	U	SMPL
P3455-01MS - 1	RW7-SP100-50-20240801MS	14618	3.2116	-	7470A	PPB	07 Aug 2024 12:13:49	U	SMPL
P3455-01MSD - 1	RW7-SP100-50-20240801MSD	15024	3.2999	-	7470A	PPB	07 Aug 2024 12:16:04	U	SMPL
P3455-03 - 1	RW7-S301A-50-20240801	569	0.1568	-	7470A	PPB	07 Aug 2024 12:18:22	U	SMPL
P3455-04 - 1	RW7-SP303-50-20240801	223	0.0815	-	7470A	PPB	07 Aug 2024 12:20:40	U	SMPL
P3455-05 - 1	RW7-SP100-90-20240801	167	0.0694	-	7470A	PPB	07 Aug 2024 12:22:56	U	SMPL
P3455-07 - 1	RW7-SP301A-90-20240801	192	0.0748	-	7470A	PPB	07 Aug 2024 12:25:12	U	SMPL
P3455-08 - 1	RW7-SP303-90-20240801	96	0.0539	-	7470A	PPB	07 Aug 2024 12:27:28	U	SMPL
P3490-01 - 1	402	244	0.0861	-	7470A	PPB	07 Aug 2024 12:29:45	U	SMPL
CCV07 - 1	CCV07	21216	4.6463	-	7470A	PPB	07 Aug 2024 12:32:01	U	SMPL
CCB07 - 1	CCB07	6	0.0343	-	7470A	PPB	07 Aug 2024 12:34:17	U	SMPL
PB162509TB - 1		-4	0.0322	-	7470A	PPB	07 Aug 2024 12:36:33	U	SMPL
P3474-11LX5 - 1		-30	0.0285	-	7470A	PPB	07 Aug 2024 12:38:48	U	SMPL
P3474-11A - 1		16656	3.6548	-	7470A	PPB	07 Aug 2024 12:41:04	U	SMPL
P3455-01LX5 - 1		-154	-0.0004	-	7470A	PPB	07 Aug 2024 12:43:19	U	SMPL
P3455-01A - 1		16849	3.6968	-	7470A	PPB	07 Aug 2024 12:45:34	U	SMPL
CCV08 - 1	CCV08	22440	4.9125	-	7470A	PPB	07 Aug 2024 12:47:50	U	SMPL
CCB08 - 1	CCB08	-243	-0.0198	-	7470A	PPB	07 Aug 2024 12:50:07	U	SMPL
							07 Aug 2024 12:53:24	U	SMPL