

LB 131894


7471B
MR 7470A
8/9/24

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40181



A= 0.0000e+000

B= 2.4634e-004

C= -2.4803e-002

Rho= 0.9996374

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.017	0.017	171	0.000	171					
0.2	0.200	0.190	-0.010	871	0.0 %	871					
2.5	2.500	2.455	-0.045	10065	0.0 %	10065					
5.0	5.000	4.966	-0.034	20259	0.0 %	20259					
7.5	7.500	7.699	0.199	31354	0.0 %	31354					
10.0	10.000	9.873	-0.127	40181	0.0 %	40181					

1
 -5
 -2
 -1
 3
 -1

LB131894 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Aug 2024 13:58:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	171	-	0.0000	7470A	PPB	07 Aug 2024 14:03:10	S	Std
0.2 - 1	50.2	871	-	0.2000	7470A	PPB	07 Aug 2024 14:05:27	S	Std
2.5 - 1	52.5	10065	-	2.5000	7470A	PPB	07 Aug 2024 14:07:44	S	Std
5.0 - 1	55	20259	-	5.0000	7470A	PPB	07 Aug 2024 14:10:01	S	Std
7.5 - 1	57.5	31354	-	7.5000	7470A	PPB	07 Aug 2024 14:12:18	S	Std
10.0 - 1	511	40181	-	10.0000	7470A	PPB	07 Aug 2024 14:14:35	S	Std
ICV64 - 1	ICV64	15695	3.8415	-	7470A	PPB	07 Aug 2024 14:17:24	U	SMPL
ICB64 - 1	ICB64	-33	-0.0329	-	7470A	PPB	07 Aug 2024 14:19:39	U	SMPL
CCV09 - 1	CCV09	20255	4.9649	-	7470A	PPB	07 Aug 2024 14:21:56	U	SMPL
CCB09 - 1	CCB09	-37	-0.0339	-	7470A	PPB	07 Aug 2024 14:24:10	U	SMPL
CRA - 1	CRA	839	0.1819	-	7470A	PPB	07 Aug 2024 14:26:27	U	SMPL
HighStd - 1	HighStd	40405	9.9286	-	7470A	PPB	07 Aug 2024 14:28:42	U	SMPL
ChkStd - 1	ChkStd	25854	6.3441	-	7470A	PPB	07 Aug 2024 14:30:57	U	SMPL
PB162560BL - 1	PBS	-143	-0.0600	-	7470A	PPB	07 Aug 2024 14:33:15	U	SMPL
PB162560BS - 1	LCSS	16086	3.9379	-	7470A	PPB	07 Aug 2024 14:35:33	U	SMPL
P3468-01 - 1	IDW-S1-01	117195	28.8452	-	7470A	PPB	07 Aug 2024 14:37:49	U	SMPL
P3474-01 - 1	WC-A1-COMP	2770	0.6576	-	7470A	PPB	07 Aug 2024 14:45:18	U	SMPL
P3474-01DUP - 1	WC-A1-COMPDUP	2564	0.6068	-	7470A	PPB	07 Aug 2024 14:47:35	U	SMPL
P3474-01MS - 1	WC-A1-COMPMS	20193	4.9496	-	7470A	PPB	07 Aug 2024 14:49:52	U	SMPL
P3474-01MSD - 1	WC-A1-COMPMSD	20097	4.9259	-	7470A	PPB	07 Aug 2024 14:55:01	U	SMPL
CCV10 - 1	CCV10	19835	4.8614	-	7470A	PPB	07 Aug 2024 14:57:16	U	SMPL
CCB10 - 1	CCB10	-8	-0.0268	-	7470A	PPB	07 Aug 2024 14:59:33	U	SMPL
P3474-04 - 1	WC-B1-COMP	2013	0.4711	-	7470A	PPB	07 Aug 2024 15:01:49	U	SMPL
P3474-07 - 1	WC-A2-COMP	2330	0.5492	-	7470A	PPB	07 Aug 2024 15:04:04	U	SMPL
P3474-10 - 1	WC-B2-COMP	1195	0.2696	-	7470A	PPB	07 Aug 2024 15:06:19	U	SMPL
P3483-01 - 1	A	2513	0.5943	-	7470A	PPB	07 Aug 2024 15:08:34	U	SMPL
P3484-01 - 1	B	6806	1.6518	-	7470A	PPB	07 Aug 2024 15:10:50	U	SMPL
P3485-01 - 1	C	3393	0.8110	-	7470A	PPB	07 Aug 2024 15:13:07	U	SMPL
P3488-01 - 1	HD-01-080624	1159	0.2607	-	7470A	PPB	07 Aug 2024 15:15:23	U	SMPL
P3488-03 - 1	HD-02-080624	1041	0.2316	-	7470A	PPB	07 Aug 2024 15:17:39	U	SMPL
P3493-01 - 1	628-A	776	0.1664	-	7470A	PPB	07 Aug 2024 15:19:57	U	SMPL
P3493-02 - 1	628-B	990	0.2191	-	7470A	PPB	07 Aug 2024 15:22:12	U	SMPL
CCV11 - 1	CCV11	19957	4.8914	-	7470A	PPB	07 Aug 2024 15:24:28	U	SMPL
CCB11 - 1	CCB11	14	-0.0214	-	7470A	PPB	07 Aug 2024 15:26:43	U	SMPL
P3474-01LX5 - 1		461	0.0888	-	7470A	PPB	07 Aug 2024 15:29:00	U	SMPL
P3468-01DLX5 - 1		25648	6.2934	-	7470A	PPB	07 Aug 2024 15:33:31	U	SMPL
P3474-01A - 1		17529	4.2933	-	7470A	PPB	07 Aug 2024 15:38:49	U	SMPL
CCV12 - 1	CCV12	20285	4.9722	-	7470A	PPB	07 Aug 2024 15:41:36	U	SMPL
CCB12 - 1	CCB12	19	-0.0201	-	7470A	PPB	07 Aug 2024 15:43:52	U	SMPL