


LB131296

SFAM01.1

INSTRUMENT ID: CV1



A= 0.0000e+000
 B= 2.3413e-004 *slope*
 C= -2.3877e-002 *y-intercept*
 Rho= 0.9998843
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>6/03</i>
0.00	0.000	-0.003	-0.003	90	0.000	90					
0.05	0.050					442	0				
0.20	0.200	0.215	0.015	1020	0.0 %	1020					
2.50	2.500	2.532	0.032	10916	0.0 %	10916					
5.00	5.000	4.986	-0.014	21398	0.0 %	21398					
7.50	7.500	7.393	-0.107	31678	0.0 %	31678					
10.0	10.000	10.077	0.077	43144	0.0 %	43144					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	90	-		0 SFAM01.1	PPB	8/7/2024 16:45	Std
	0.2 S01	1020	-		0.2 SFAM01.1	PPB	8/7/2024 16:48	Std
	2.5 S02	10916	-		2.5 SFAM01.1	PPB	8/7/2024 16:50	Std
	5 S03	21398	-		5 SFAM01.1	PPB	8/7/2024 16:55	Std
	7.5 S04	31678	-		7.5 SFAM01.1	PPB	8/7/2024 16:57	Std
	10 S05	43144	-		10 SFAM01.1	PPB	8/7/2024 16:59	Std
ICV066	ICV066	17789	4.141 -		SFAM01.1	PPB	8/7/2024 17:02	SMPL
ICB066	ICB066	-44	-0.0342 -		SFAM01.1	PPB	8/7/2024 17:04	SMPL
CCV013	CCV013	21671	5.0499 -		SFAM01.1	PPB	8/7/2024 17:07	SMPL
CCB013	CCB013	-56	-0.037 -		SFAM01.1	PPB	8/7/2024 17:09	SMPL
PB162534BL	PBW534	61	-0.0096 -		SFAM01.1	PPB	8/7/2024 17:11	SMPL
P3428-01	MH1Y26	2109	0.4699 -		SFAM01.1	PPB	8/7/2024 17:13	SMPL
P3428-02	MH1Y28	140	0.0089 -		SFAM01.1	PPB	8/7/2024 17:16	SMPL
P3428-03	MH1Y30	136	0.008 -		SFAM01.1	PPB	8/7/2024 17:18	SMPL
P3428-04	MH1Y32	151	0.0115 -		SFAM01.1	PPB	8/7/2024 17:20	SMPL
P3428-05	MH1Y34	183	0.019 -		SFAM01.1	PPB	8/7/2024 17:23	SMPL
P3428-06	MH1Y36	1264	0.2721 -		SFAM01.1	PPB	8/7/2024 17:25	SMPL
P3428-07	MH1Y38	2458	0.5516 -		SFAM01.1	PPB	8/7/2024 17:27	SMPL
P3428-08	MH1Y40	4497	1.029 -		SFAM01.1	PPB	8/7/2024 17:29	SMPL
P3428-09	MH1Y42	1657	0.3641 -		SFAM01.1	PPB	8/7/2024 17:32	SMPL
P3428-10	MH1Y42D	1584	0.347 -		SFAM01.1	PPB	8/7/2024 17:34	SMPL
P3428-11	MH1Y42S	6492	1.4961 -		SFAM01.1	PPB	8/7/2024 17:36	SMPL
PB162566BL	PBW566	11	-0.0213 -		SFAM01.1	PPB	8/7/2024 17:38	SMPL
P3427-01	MH1X87	230	0.03 -		SFAM01.1	PPB	8/7/2024 17:41	SMPL
P3427-02	MH1X89	2877	0.6497 -		SFAM01.1	PPB	8/7/2024 17:43	SMPL
P3427-03	MH1X91	298	0.0459 -		SFAM01.1	PPB	8/7/2024 17:45	SMPL
P3427-04	MH1X93	669	0.1328 -		SFAM01.1	PPB	8/7/2024 17:47	SMPL
P3427-05	MH1X95	185	0.0194 -		SFAM01.1	PPB	8/7/2024 17:50	SMPL
P3427-06	MH1X97	2725	0.6141 -		SFAM01.1	PPB	8/7/2024 17:52	SMPL
P3427-07	MH1X99	230	0.03 -		SFAM01.1	PPB	8/7/2024 17:54	SMPL
CCV014	CCV014	21783	5.0761 -		SFAM01.1	PPB	8/7/2024 17:56	SMPL
CCB014	CCB014	-115	-0.0508 -		SFAM01.1	PPB	8/7/2024 17:59	SMPL
P3427-08	MH1Y02	153	0.0119 -		SFAM01.1	PPB	8/7/2024 18:01	SMPL
P3427-09	MH1Y04	57	-0.0105 -		SFAM01.1	PPB	8/7/2024 18:03	SMPL
P3427-10	MH1Y06	176	0.0173 -		SFAM01.1	PPB	8/7/2024 18:06	SMPL
P3427-11	MH1Y08	84	-0.0042 -		SFAM01.1	PPB	8/7/2024 18:08	SMPL
P3427-12	MH1Y10	325	0.0522 -		SFAM01.1	PPB	8/7/2024 18:10	SMPL
P3427-13	MH1Y12	2475	0.5556 -		SFAM01.1	PPB	8/7/2024 18:12	SMPL
P3427-14	MH1Y14	213	0.026 -		SFAM01.1	PPB	8/7/2024 18:15	SMPL
P3427-15	MH1Y16	54	-0.0112 -		SFAM01.1	PPB	8/7/2024 18:17	SMPL
P3427-16	MH1Y18	70	-0.0075 -		SFAM01.1	PPB	8/7/2024 18:19	SMPL
P3427-17	MH1Y20	355	0.0592 -		SFAM01.1	PPB	8/7/2024 18:21	SMPL
P3427-18	MH1Y22	227	0.0293 -		SFAM01.1	PPB	8/7/2024 18:24	SMPL
P3427-19	MH1Y24	615	0.1201 -		SFAM01.1	PPB	8/7/2024 18:26	SMPL
P3427-20	MH1Y24D	529	0.1 -		SFAM01.1	PPB	8/7/2024 18:28	SMPL
P3427-21	MH1Y24S	5493	1.2622 -		SFAM01.1	PPB	8/7/2024 18:34	SMPL

LB131896

INSTRUMENT ID : CV1

PB162441TB	LEB441	49	-0.0124 -	SFAM01.1	PPB	8/7/2024 18:36	SMPL
PB162497TB	LEB497	78	-0.0056 -	SFAM01.1	PPB	8/7/2024 18:39	SMPL
CCV015	CCV015	21644	5.0436 -	SFAM01.1	PPB	8/7/2024 18:41	SMPL
CCB015	CCB015	-111	-0.0499 -	SFAM01.1	PPB	8/7/2024 18:43	SMPL