

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB131897

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	08/07/2024	11:10
2	P3487-01		1	8	60.00	140.0	139.7	08/07/2024	11:40
3	P3487-02		1	8	59.00	138.2	137.9	08/07/2024	12:10
4	P3487-03		1	8	59.00	138.2	137.9	08/07/2024	12:40
5	P3487-05		1	11	100.00	>212.0	>212.0	08/07/2024	13:10
6	P3487-05DUP		1	11	100.00	>212.0	>212.0	08/07/2024	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6131897

WorkList Name : fp-s-8-7

WorkList ID : 182421

Department : Wet-Chemistry

Date : 08-07-2024 08:49:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3487-01	0726-A	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	08/06/2024	1010B
P3487-02	0726-B	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	08/06/2024	1010B
P3487-03	0726-C	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	08/06/2024	1010B
P3487-05	0723-B	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	08/06/2024	1010B

Date/Time 08/07/2024 11:00
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC

Date/Time 08/07/2024 14:20
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC