


LB131903

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.016	0.016	25	0.000	25	0				—
0.2	0.200	0.190	-0.010	746	0.0 %	746					—5
2.5	2.500	2.400	-0.100	9886	0.0 %	9886					—4
5.0	5.000	5.171	0.171	21347	0.0 %	21347					3
7.5	7.500	7.449	-0.051	30770	0.0 %	30770					1
10.0	10.000	9.974	-0.026	41217	0.0 %	41217					0

~~LB131903~~ INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Aug 2024 11:55:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000	7470A	PPB	08 Aug 2024 13:35:08	S	Std
0.2 - 1	50.2	746	-	0.2000	7470A	PPB	08 Aug 2024 13:37:25	S	Std
2.5 - 1	52.5	9886	-	2.5000	7470A	PPB	08 Aug 2024 13:39:42	S	Std
5.0 - 1	55	21347	-	5.0000	7470A	PPB	08 Aug 2024 13:41:59	S	Std
7.5 - 1	57.5	30770	-	7.5000	7470A	PPB	08 Aug 2024 13:44:15	S	Std
10.0 - 1	510	41217	-	10.0000	7470A	PPB	08 Aug 2024 13:46:33	S	Std
ICV67 - 1	ICV67	16971	4.1128	-	7470A	PPB	08 Aug 2024 13:54:23	U	SMPL
ICB67 - 1	ICB67	-152	-0.0266	-	7470A	PPB	08 Aug 2024 13:56:38	U	SMPL
CCV16 - 1	CCV16	20930	5.0698	-	7470A	PPB	08 Aug 2024 13:58:54	U	SMPL
CCB16 - 1	CCB16	-126	-0.0203	-	7470A	PPB	08 Aug 2024 14:01:09	U	SMPL
CRA - 1	CRA	765	0.1950	-	7470A	PPB	08 Aug 2024 14:03:26	U	SMPL
HighStd - 1	HighStd	37359	9.0415	-	7470A	PPB	08 Aug 2024 14:09:25	U	SMPL
ChkStd - 1	ChkStd	27135	6.5699	-	7470A	PPB	08 Aug 2024 14:11:40	U	SMPL
PB162581BL - 1	PBW	-298	-0.0619	-	7470A	PPB	08 Aug 2024 14:13:58	U	SMPL
PB162581BS - 1	LCSW	14997	3.6356	-	7470A	PPB	08 Aug 2024 14:16:16	U	SMPL
P3464-02 - 1	TOPSOIL	-120	-0.0189	-	7470A	PPB	08 Aug 2024 14:18:32	U	SMPL
P3476-01 - 1	MUL04	61	0.0249	-	7470A	PPB	08 Aug 2024 14:20:50	U	SMPL
P3476-03 - 1	MUL05	98	0.0338	-	7470A	PPB	08 Aug 2024 14:23:06	U	SMPL
P3476-05 - 1	MUL06	126	0.0406	-	7470A	PPB	08 Aug 2024 14:25:23	U	SMPL
P3476-07 - 1	MUL07	104	0.0353	-	7470A	PPB	08 Aug 2024 14:27:38	U	SMPL
CCV17 - 1	CCV17	21009	5.0889	-	7470A	PPB	08 Aug 2024 14:29:54	U	SMPL
CCB17 - 1	CCB17	-175	-0.0322	-	7470A	PPB	08 Aug 2024 14:32:09	U	SMPL
P3479-01 - 1	TS08	19	0.0147	-	7470A	PPB	08 Aug 2024 14:34:26	U	SMPL
P3479-03 - 1	TS09	71	0.0273	-	7470A	PPB	08 Aug 2024 14:36:41	U	SMPL
P3479-05 - 1	TS10	17	0.0142	-	7470A	PPB	08 Aug 2024 14:38:57	U	SMPL
P3479-07 - 1	TS11	64	0.0256	-	7470A	PPB	08 Aug 2024 14:41:12	U	SMPL
PB162582BL - 1	PBW	28	0.0169	-	7470A	PPB	08 Aug 2024 14:43:28	U	SMPL
PB162582BS - 1	LCSW	16462	3.9897	-	7470A	PPB	08 Aug 2024 14:45:44	U	SMPL
P3491-01 - 1	72-11936	-136	-0.0228	-	7470A	PPB	08 Aug 2024 14:48:00	U	SMPL
P3491-02 - 1	SOM-24-00087	-18	0.0058	-	7470A	PPB	08 Aug 2024 14:50:18	U	SMPL
P3491-03 - 1	COMP-1	22	0.0154	-	7470A	PPB	08 Aug 2024 14:52:35	U	SMPL
P3491-04 - 1	COMP-2	53	0.0229	-	7470A	PPB	08 Aug 2024 14:54:51	U	SMPL
CCV18 - 1	CCV18	21508	5.2096	-	7470A	PPB	08 Aug 2024 14:57:06	U	SMPL
CCB18 - 1	CCB18	-142	-0.0242	-	7470A	PPB	08 Aug 2024 14:59:22	U	SMPL
P3491-05 - 1	SOM-24-00094	-17	0.0060	-	7470A	PPB	08 Aug 2024 15:01:39	U	SMPL
P3506-02 - 1	RT4534	10	0.0125	-	7470A	PPB	08 Aug 2024 15:03:54	U	SMPL
P3506-02DUP - 1	RT4534DUP	14	0.0135	-	7470A	PPB	08 Aug 2024 15:06:09	U	SMPL
P3506-02MS - 1	RT4534MS	17351	4.2046	-	7470A	PPB	08 Aug 2024 15:08:25	U	SMPL
P3506-02MSD - 1	RT4534MSD	14968	3.6286	-	7470A	PPB	08 Aug 2024 15:10:41	U	SMPL
PB162441TB - 1	PB162562TB	-139	-0.0235	-	7470A	PPB	08 Aug 2024 15:12:58	U	SMPL
PB162497TB - 1	PB162563TB	25	0.0162	-	7470A	PPB	08 Aug 2024 15:15:16	U	SMPL
P3478-01 - 1	TS04	206	0.0599	-	7470A	PPB	08 Aug 2024 15:17:32	U	SMPL
P3478-01DUP - 1	TS04DUP	201	0.0587	-	7470A	PPB	08 Aug 2024 15:19:49	U	SMPL
P3478-02 - 1	P3478-01MS	17301	4.1925	-	7470A	PPB	08 Aug 2024 15:26:28	U	SMPL
CCV19 - 1	CCV19	19251	4.6639	-	7470A	PPB	08 Aug 2024 15:28:44	U	SMPL
CCB19 - 1	CCB19	-190	-0.0358	-	7470A	PPB	08 Aug 2024 15:31:01	U	SMPL
P3478-03 - 1	P3478-01MSD	17174	4.1618	-	7470A	PPB	08 Aug 2024 15:33:18	U	SMPL
P3478-05 - 1	TS05	-156	-0.0276	-	7470A	PPB	08 Aug 2024 15:35:33	U	SMPL
P3478-07 - 1	TS06	-11	0.0075	-	7470A	PPB	08 Aug 2024 15:37:50	U	SMPL
P3478-09 - 1	TS07	28	0.0169	-	7470A	PPB	08 Aug 2024 15:40:06	U	SMPL
P3506-02LX5 - 1		14	0.0135	-	7470A	PPB	08 Aug 2024 15:42:22	U	SMPL
P3506-02A - 1		17162	4.1589	-	7470A	PPB	08 Aug 2024 15:44:38	U	SMPL
P3478-01LX5 - 1		-113	-0.0172	-	7470A	PPB	08 Aug 2024 15:46:55	U	SMPL
P3478-01A - 1		16090	3.8998	-	7470A	PPB	08 Aug 2024 15:49:13	U	SMPL
PB162563TB - 1	PB162581TB	-122	-0.0194	-	7470A	PPB	08 Aug 2024 15:51:30	U	SMPL
CCV20 - 1	CCV20	19952	4.8334	-	7470A	PPB	08 Aug 2024 15:53:47	U	SMPL
CCB20 - 1	CCB20	-165	-0.0298	-	7470A	PPB	08 Aug 2024 15:56:02	U	SMPL

MB

08/08/2024