

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131911

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/08/2024	14:30
2	CAL2	1	Water	NA	NA	20.2	7.00	08/08/2024	14:31
3	CAL3	1	Water	NA	NA	20.3	10.02	08/08/2024	14:32
4	ICV	1	Water	NA	NA	20.3	7.02	08/08/2024	14:35
5	CCV1	1	Water	NA	NA	20.3	2.01	08/08/2024	14:40
6	P3530-02	1	Water	NA	NA	20.6	11.85	08/08/2024	14:50
7	P3530-02DUP	1	Water	NA	NA	20.7	11.86	08/08/2024	14:51
8	P3530-03	1	Water	NA	NA	20.7	8.66	08/08/2024	15:00
9	CCV2	1	Water	NA	NA	20.3	12.02	08/08/2024	15:05

WORKLIST(Hardcopy Internal Chain)

8/13/11

WorkList Name : ph p3530

WorkList ID : 182464

Department : Wet-Chemistry

Date : 08-08-2024 14:14:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3530-02	WC-A-04-202408	Water	pH	Cool 4 deg C	AECO02	D21	08/08/2024	9040C
P3530-03	WC-A-05-202408	Water	pH	Cool 4 deg C	AECO02	D21	08/08/2024	9040C

Date/Time

08/08/24 14120

Raw Sample Received by:

70 (WC)

Raw Sample Relinquished by:

CP

Date/Time

08/08/24

Raw Sample Received by:

CP

Raw Sample Relinquished by:

JP