

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131912

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/08/2024	14:30
2	CAL2	1	Water	NA	NA	20.2	7.00	08/08/2024	14:31
3	CAL3	1	Water	NA	NA	20.3	10.02	08/08/2024	14:32
4	ICV	1	Water	NA	NA	20.3	7.02	08/08/2024	14:35
5	CCV1	1	Water	NA	NA	20.4	2.01	08/08/2024	15:10
6	P3506-02	1	Solid	20.02	20	22.2	5.32	08/08/2024	15:15
7	P3526-02	1	Solid	20.03	20	20.3	8.53	08/08/2024	15:20
8	P3526-02DUP	1	Solid	20.04	20	20.4	8.54	08/08/2024	15:21
9	CCV2	1	Water	NA	NA	20.3	12.02	08/08/2024	15:25

8/13/12

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p3526 WorkList ID : 182465 Department : Wet-Chemistry Date : 08-08-2024 14:15:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3506-02	RT4534	Solid	Corrosivity	Cool 4 deg C	PSEG03	D21	08/07/2024	9045D
P3526-02	SP-1-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	D21	08/07/2024	9045D

Date/Time 08/08/24 14:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 08/08/24 17:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]