

LB131914


SFAM01.1

INSTRUMENT ID: CVL

Linear

 μ Abs.:

42687



A= 0.0000e+000

B= 2.3660e-004 *slope*C= -4.1700e-002 *intercept*

Rho= 0.9999049

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.013	-0.013	121	0.000	121					=
0.05	0.050					442	0				1
0.20	0.200	0.199	-0.001	1019	0.0 %	1019					2
2.50	2.500	2.554	0.054	10973	0.0 %	10973					0
5.00	5.000	4.996	-0.004	21293	0.0 %	21293					1
7.50	7.500	7.405	-0.095	31474	0.0 %	31474					1
10.0	10.000	10.058	0.058	42687	0.0 %	42687					1

LB131914

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	121	-		0 SFAM01.1	PPB	8/8/2024 16:21	Std
	0.2 S01	1019	-		0.2 SFAM01.1	PPB	8/8/2024 16:23	Std
	2.5 S02	10973	-		2.5 SFAM01.1	PPB	8/8/2024 16:25	Std
	5 S03	21293	-		5 SFAM01.1	PPB	8/8/2024 16:28	Std
	7.5 S04	31474	-		7.5 SFAM01.1	PPB	8/8/2024 16:30	Std
	10 S05	42687	-		10 SFAM01.1	PPB	8/8/2024 16:32	Std
ICV068	ICV068	17108	4.006 -		SFAM01.1	PPB	8/8/2024 16:35	SMPL
ICB068	ICB068	-8	-0.0436 -		SFAM01.1	PPB	8/8/2024 16:37	SMPL
CCV021	CCV021	21475	5.0393 -		SFAM01.1	PPB	8/8/2024 16:40	SMPL
CCB021	CCB021	-66	-0.0573 -		SFAM01.1	PPB	8/8/2024 16:42	SMPL
PB162592BL	PBW592	11	-0.0391 -		SFAM01.1	PPB	8/8/2024 16:44	SMPL
P3499-01	E08L9	457	0.0664 -		SFAM01.1	PPB	8/8/2024 16:51	SMPL
P3499-02	E08L9D	347	0.0404 -		SFAM01.1	PPB	8/8/2024 16:53	SMPL
P3499-03	E08L9S	4481	1.0185 -		SFAM01.1	PPB	8/8/2024 16:56	SMPL
P3499-04	E08M2	-16	-0.0455 -		SFAM01.1	PPB	8/8/2024 16:58	SMPL
P3500-01	E08M0	8	-0.0398 -		SFAM01.1	PPB	8/8/2024 17:00	SMPL
P3500-02	E08MOD	24	-0.036 -		SFAM01.1	PPB	8/8/2024 17:03	SMPL
P3500-03	E08MOS	4240	0.9615 -		SFAM01.1	PPB	8/8/2024 17:05	SMPL
P3500-04	E08M3	-9	-0.0438 -		SFAM01.1	PPB	8/8/2024 17:07	SMPL
CCV022	CCV022	20494	4.8072 -		SFAM01.1	PPB	8/8/2024 17:09	SMPL
CCB022	CCB022	-119	-0.0699 -		SFAM01.1	PPB	8/8/2024 17:12	SMPL