

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131920

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/09/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.01	08/09/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	08/09/2024	08:12
4	ICV	1	Water	NA	NA	20.6	7.02	08/09/2024	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	08/09/2024	08:20
6	P3534-05	1	Water	NA	NA	20.1	5.67	08/09/2024	08:25
7	P3534-05DUP	1	Water	NA	NA	20.2	5.68	08/09/2024	08:26
8	P3534-06	1	Water	NA	NA	20.2	5.52	08/09/2024	08:30
9	CCV2	1	Water	NA	NA	20.2	12.02	08/09/2024	08:35

UP 131920

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p3534 WorkList ID : 182482 Department : Wet-Chemistry Date : 08-09-2024 07:54:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3534-05	PIPE-BLAST-GRIT(FB)	Water	pH	Cool 4 deg C	PSEG04	D11	08/08/2024	9040C
P3534-06	HP-IP(FB)	Water	pH	Cool 4 deg C	PSEG04	D11	08/08/2024	9040C

Date/Time 08/09/24 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

10:00
Date/Time 08/09/24 10:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]