

Analytical Summary Report

Analysis Method: 9045D
Parameter: Corrosivity
Run Number: LB131921
BalanceID: WC SC-4

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 98.6
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/09/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.01	08/09/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	08/09/2024	08:12
4	ICV	1	Water	NA	NA	20.6	7.02	08/09/2024	08:15
5	CCV1	1	Water	NA	NA	20.4	2.02	08/09/2024	08:40
6	P3534-01	1	Solid	20.02	20	20.2	5.49	08/09/2024	08:45
7	P3534-01DUP	1	Solid	20.01	20	20.3	5.50	08/09/2024	08:46
8	P3534-03	1	Solid	20.03	20	20.4	6.06	08/09/2024	08:50
9	CCV2	1	Water	NA	NA	20.5	12.02	08/09/2024	08:55

WORKLIST(Hardcopy Internal Chain)

WS13192)

WorkList Name : corrosivity p3534 **WorkList ID :** 182483 **Department :** Wet-Chemistry **Date :** 08-09-2024 07:55:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3534-01	PIPE-BLAST-GRIT	Solid	Corrosivity	Cool 4 deg C	PSEG04	D21	08/08/2024	9045D
P3534-03	HP-IP	Solid	Corrosivity	Cool 4 deg C	PSEG04	D21	08/08/2024	9045D

Date/Time 08/09/24 08:00
Raw Sample Received by: JL WGC
Raw Sample Relinquished by: [Signature]

Date/Time 08/09/24 10:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]