

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131924

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/09/2024	10:35
2	P3487-04		1	11	100.00	>212.0	>212.0	08/09/2024	11:00
3	P3487-04DUP		1	11	100.00	>212.0	>212.0	08/09/2024	11:30
4	P3530-02		1	14	100.00	>212.0	>212.0	08/09/2024	12:00
5	P3530-03		1	11	100.00	>212.0	>212.0	08/09/2024	12:30
6	P3532-01		1	13	100.00	>212.0	>212.0	08/09/2024	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66131924

WorkList Name : FP8-8

WorkList ID : 182477

Department : Wet-Chemistry

Date : 08-08-2024 18:52:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3487-04	584	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/06/2024	1010B
P3530-02	WC-A-04-202408	Water	Flash Point	Cool 4 deg C	AECO02	D21	08/08/2024	1010B
P3530-03	WC-A-05-202408	Water	Flash Point	Cool 4 deg C	AECO02	D21	08/08/2024	1010B
P3532-01	806-C	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/08/2024	1010B

Date/Time 08/09/2024 10.25
Raw Sample Received by: RM CW
Raw Sample Relinquished by: RM CW

Date/Time 08/09/2024 13.40
Raw Sample Received by: RM CW
Raw Sample Relinquished by: RM CW