


LB131927

7471B

INSTRUMENT ID: CV 1

 μ Abs.:

46267



Linear

A= 0.0000e+000

B= 2.1419e-004

C= -4.9905e-003

Rho= 0.9995591

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.018	0.018	108	0.000	108					-6
0.2	0.200	0.189	-0.011	906	0.0 %	906					-6
2.5	2.500	2.355	-0.145	11018	0.0 %	11018					4
5.0	5.000	5.201	0.201	24307	0.0 %	24307					0
7.5	7.500	7.531	0.031	35184	0.0 %	35184					1
10.0	10.000	9.905	-0.095	46267	0.0 %	46267					

LB131927 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Aug 2024 14:34:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	108	-	0.0000	7471B	PPB	09 Aug 2024 14:34:33	S	Std
0.2 - 1	50.2	906	-	0.2000	7471B	PPB	09 Aug 2024 14:36:50	S	Std
2.5 - 1	52.5	11018	-	2.5000	7471B	PPB	09 Aug 2024 14:39:07	S	Std
5.0 - 1	55	24307	-	5.0000	7471B	PPB	09 Aug 2024 14:44:08	S	Std
7.5 - 1	57.5	35184	-	7.5000	7471B	PPB	09 Aug 2024 14:48:58	S	Std
10.0 - 1	59	46267	-	10.0000	7471B	PPB	09 Aug 2024 14:51:15	S	Std
ICV69 - 1	ICV69	18470	3.9512	-	7471B	PPB	09 Aug 2024 14:54:20	U	SMPL
ICB69 - 1	ICB69	-194	-0.0465	-	7471B	PPB	09 Aug 2024 14:56:36	U	SMPL
CCV23 - 1	CCV23	24102	5.1575	-	7471B	PPB	09 Aug 2024 14:58:53	U	SMPL
CCB23 - 1	CCB23	-210	-0.0500	-	7471B	PPB	09 Aug 2024 15:01:08	U	SMPL
CRA - 1	CRA	857	0.1786	-	7471B	PPB	09 Aug 2024 15:03:25	U	SMPL
HighStd - 1	HighStd	45423	9.7244	-	7471B	PPB	09 Aug 2024 15:05:40	U	SMPL
ChkStd - 1	ChkStd	32765	7.0131	-	7471B	PPB	09 Aug 2024 15:07:56	U	SMPL
PB162620BL - 1	PBS	-310	-0.0714	-	7471B	PPB	09 Aug 2024 15:10:19	U	SMPL
PB162620BS - 1	LCSS	18792	4.0201	-	7471B	PPB	09 Aug 2024 15:12:38	U	SMPL
P3506-01 - 1	RT4534	7812	1.6683	-	7471B	PPB	09 Aug 2024 15:14:55	U	SMPL
P3507-01 - 1	EO-03-080724	2521	0.5350	-	7471B	PPB	09 Aug 2024 15:17:13	U	SMPL
P3509-01 - 1	OR-02-080724	3328	0.7078	-	7471B	PPB	09 Aug 2024 15:19:33	U	SMPL
P3519-01 - 1	OU4-TS-16-080724	4186	0.8916	-	7471B	PPB	09 Aug 2024 15:21:50	U	SMPL
P3519-03 - 1	OU4-TS-17-080724	4264	0.9083	-	7471B	PPB	09 Aug 2024 15:24:06	U	SMPL
CCV24 - 1	CCV24	23681	5.0673	-	7471B	PPB	09 Aug 2024 15:26:23	U	SMPL
CCB24 - 1	CCB24	-172	-0.0418	-	7471B	PPB	09 Aug 2024 15:28:39	U	SMPL
P3519-05 - 1	OU4-TS-18-080724	4686	0.9987	-	7471B	PPB	09 Aug 2024 15:30:57	U	SMPL
P3519-07 - 1	OU4-TS-19-080724	4301	0.9163	-	7471B	PPB	09 Aug 2024 15:33:12	U	SMPL
P3519-09 - 1	OU4-TS-20-080724	3684	0.7841	-	7471B	PPB	09 Aug 2024 15:35:29	U	SMPL
P3519-11 - 1	OU4-TS-21-080724	5184	1.1054	-	7471B	PPB	09 Aug 2024 15:37:46	U	SMPL
P3519-13 - 1	OU4-TS-22-080724	4778	1.0184	-	7471B	PPB	09 Aug 2024 15:40:02	U	SMPL
P3519-15 - 1	OU4-TS-23-080724	5780	1.2331	-	7471B	PPB	09 Aug 2024 15:42:20	U	SMPL
P3519-17 - 1	OU4-CF-01-080724	87	0.0136	-	7471B	PPB	09 Aug 2024 15:44:38	U	SMPL
P3519-19 - 1	OU4-CF-02-080724	163	0.0299	-	7471B	PPB	09 Aug 2024 15:46:56	U	SMPL
P3519-21 - 1	OU4-CF-03-080724	79	0.0119	-	7471B	PPB	09 Aug 2024 15:49:13	U	SMPL
P3519-23 - 1	OU4-CF-04-080724	166	0.0306	-	7471B	PPB	09 Aug 2024 15:51:28	U	SMPL
CCV25 - 1	CCV25	23644	5.0594	-	7471B	PPB	09 Aug 2024 15:53:43	U	SMPL
CCB25 - 1	CCB25	-160	-0.0393	-	7471B	PPB	09 Aug 2024 15:55:59	U	SMPL
P3519-25 - 1	OU4-CF-05-080724	501	0.1023	-	7471B	PPB	09 Aug 2024 15:58:17	U	SMPL
P3519-27 - 1	OU4-CF-06-080724	125	0.0218	-	7471B	PPB	09 Aug 2024 16:00:32	U	SMPL
P3519-29 - 1	OU4-CF-07-080724	306	0.0606	-	7471B	PPB	09 Aug 2024 16:02:48	U	SMPL
P3519-29DUP - 1	OU4-CF-07-080724DUP	274	0.0537	-	7471B	PPB	09 Aug 2024 16:05:03	U	SMPL
P3519-29MS - 1	OU4-CF-07-080724MS	19124	4.0913	-	7471B	PPB	09 Aug 2024 16:07:19	U	SMPL
P3519-29MSD - 1	OU4-CF-07-080724MSD	18723	4.0054	-	7471B	PPB	09 Aug 2024 16:09:34	U	SMPL
P3519-29LX5 - 1		-132	-0.0333	-	7471B	PPB	09 Aug 2024 16:11:53	U	SMPL
P3519-29A - 1		18861	4.0349	-	7471B	PPB	09 Aug 2024 16:14:13	U	SMPL
CCV26 - 1	CCV26	23586	5.0470	-	7471B	PPB	09 Aug 2024 16:16:30	U	SMPL
CCB26 - 1	CCB26	-130	-0.0328	-	7471B	PPB	09 Aug 2024 16:18:48	U	SMPL