

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131933

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/09/2024	14:45
2	CAL2	1	Water	NA	NA	20.2	7.00	08/09/2024	14:46
3	CAL3	1	Water	NA	NA	20.2	10.02	08/09/2024	14:50
4	ICV	1	Water	NA	NA	20.3	7.00	08/09/2024	14:52
5	CCV1	1	Water	NA	NA	20.3	2.02	08/09/2024	15:20
6	P3545-02	1	Solid	20.02	20	23.1	5.14	08/09/2024	15:25
7	P3545-02DUP	1	Solid	20.03	20	23.2	5.15	08/09/2024	15:26
8	CCV2	1	Water	NA	NA	20.3	12.02	08/09/2024	15:30

WORKLIST(Hardcopy Internal Chain)

✓ 131933

WorkList Name : PH S P3545

WorkList ID : 182506

Department : Wet-Chemistry

Date : 08-09-2024 14:36:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3545-02	BUR-1296	Solid	pH	Cool 4 deg C	PSEG03	D21	08/09/2024	9045D

Date/Time 08/09/24 14:40
Raw Sample Received by: SD Cully
Raw Sample Relinquished by: JWS

Date/Time 08/09/24 17:30
Raw Sample Received by: JWS
Raw Sample Relinquished by: JWS