

LB131948


INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	15	-		0 SFAM01.1	PPB	8/12/2024 12:51	Std
	0.2 S01	771	-		0.2 SFAM01.1	PPB	8/12/2024 12:53	Std
	2.5 S02	10516	-		2.5 SFAM01.1	PPB	8/12/2024 12:56	Std
	5 S03	20910	-		5 SFAM01.1	PPB	8/12/2024 12:58	Std
	7.5 S04	30495	-		7.5 SFAM01.1	PPB	8/12/2024 13:00	Std
	10 S05	41177	-		10 SFAM01.1	PPB	8/12/2024 13:05	Std
ICV071	ICV071	16032	3.8876 -		SFAM01.1	PPB	8/12/2024 13:08	SMPL
ICB071	ICB071	-158	-0.0581 -		SFAM01.1	PPB	8/12/2024 13:10	SMPL
CCV033	CCV033	20109	4.8812 -		SFAM01.1	PPB	8/12/2024 13:12	SMPL
CCB033	CCB033	-155	-0.0574 -		SFAM01.1	PPB	8/12/2024 13:15	SMPL
PB162676BL	PBW676	-1	-0.0199 -		SFAM01.1	PPB	8/12/2024 13:17	SMPL
P3528-01	MC0AA0	31	-0.0121 -		SFAM01.1	PPB	8/12/2024 13:19	SMPL
P3528-02	MC0AA0D	17	-0.0155 -		SFAM01.1	PPB	8/12/2024 13:21	SMPL
P3528-03	MC0AA0S	3735	0.8906 -		SFAM01.1	PPB	8/12/2024 13:24	SMPL
P3528-04	MC0AA2	155	0.0182 -		SFAM01.1	PPB	8/12/2024 13:26	SMPL
P3553-01	E08M6	37	-0.0106 -		SFAM01.1	PPB	8/12/2024 13:28	SMPL
P3553-02	E08M6D	34	-0.0113 -		SFAM01.1	PPB	8/12/2024 13:31	SMPL
P3553-03	E08M6S	4340	1.0381 -		SFAM01.1	PPB	8/12/2024 13:33	SMPL
P3554-01	E08M7	-48	-0.0313 -		SFAM01.1	PPB	8/12/2024 13:35	SMPL
P3554-02	E08M7D	16	-0.0157 -		SFAM01.1	PPB	8/12/2024 13:37	SMPL
P3554-03	E08M7S	4149	0.9915 -		SFAM01.1	PPB	8/12/2024 13:40	SMPL
CCV034	CCV034	19795	4.8046 -		SFAM01.1	PPB	8/12/2024 13:42	SMPL
CCB034	CCB034	-169	-0.0608 -		SFAM01.1	PPB	8/12/2024 13:44	SMPL

LB131948

SFAM01.1 INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.4371e-004 *slope*C= -1.9611e-002 *y-intercept*

Rho= 0.9998966

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>σ/0.0</i>
0.00	0.000	-0.016	-0.016	15	0.000	15					<i>=</i>
0.05	0.050					442	0				<i>-16</i>
0.20	0.200	0.168	-0.032	771	0.0 %	771					<i>2</i>
2.50	2.500	2.543	0.043	10516	0.0 %	10516					<i>2</i>
5.00	5.000	5.076	0.076	20910	0.0 %	20910					<i>-1</i>
7.50	7.500	7.412	-0.088	30495	0.0 %	30495					<i>0</i>
10.0	10.000	10.016	0.016	41177	0.0 %	41177					