

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131952

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/12/2024	14:00
2	CAL2	1	Water	NA	NA	20.3	7.01	08/12/2024	14:01
3	CAL3	1	Water	NA	NA	20.2	10.02	08/12/2024	14:05
4	ICV	1	Water	NA	NA	20.3	7.00	08/12/2024	14:10
5	CCV1	1	Water	NA	NA	20.3	2.01	08/12/2024	14:11
6	P3574-02	1	Water	NA	NA	20.4	7.37	08/12/2024	14:15
7	P3575-01	1	Water	NA	NA	24.4	7.31	08/12/2024	14:22
8	P3576-01	1	Water	NA	NA	23.3	6.81	08/12/2024	14:27
9	P3576-02	1	Water	NA	NA	23.4	7.21	08/12/2024	14:33
10	P3576-03	1	Water	NA	NA	23.4	7.02	08/12/2024	14:37
11	P3576-04	1	Water	NA	NA	23.1	7.42	08/12/2024	14:45
12	P3576-06	1	Water	NA	NA	22.7	7.72	08/12/2024	14:55
13	P3576-06DUP	1	Water	NA	NA	22.8	7.73	08/12/2024	14:56
14	CCV2	1	Water	NA	NA	20.4	12.02	08/12/2024	15:00

WORKLIST(Hardcopy Internal Chain)

VB 131952

WorkList Name : PH P3574

WorkList ID : 182566

Department : Wet-Chemistry

Date : 08-12-2024 13:46:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3574-02	WC-W-06-202408	Water	pH	Cool 4 deg C	AECO02	D21	08/12/2024	9040C
P3575-01	50350	Water	pH	Cool 4 deg C	PSEG03	D21	08/12/2024	9040C
P3576-01	801-A	Water	pH	Cool 4 deg C	PSEG03	D21	08/12/2024	9040C
P3576-02	801-B	Water	pH	Cool 4 deg C	PSEG03	D21	08/12/2024	9040C
P3576-03	801-C	Water	pH	Cool 4 deg C	PSEG03	D21	08/12/2024	9040C
P3576-04	801-D	Water	pH	Cool 4 deg C	PSEG03	D21	08/12/2024	9040C
P3576-06	NWB-2008	Water	pH	Cool 4 deg C	PSEG03	D21	08/12/2024	9040C

Date/Time 08/12/24 13:55

Raw Sample Received by: 786001

Raw Sample Relinquished by: 786001

Date/Time 08/12/24

Raw Sample Received by: 786001

Raw Sample Relinquished by: 786001