



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 8/13/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:25
In Date: 08/12/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 08/13/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131955

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3560-01	E1HX7	1	1.18	8.75	9.93	8.32	81.6	
P3560-02	E1HX8	2	1.18	8.54	9.72	8.02	80.1	
P3560-03	E1HX9	3	1.18	8.71	9.89	8.66	85.9	
P3560-04	E1HY0	4	1.15	8.80	9.95	8.48	83.3	
P3560-05	E1HY1	5	1.18	8.62	9.8	8.01	79.2	
P3560-06	E1HY2	6	1.12	8.59	9.71	8.02	80.3	
P3560-07	E1HY3	7	1.16	8.72	9.88	8.22	81.0	
P3560-08	E1HY4	8	1.12	8.72	9.84	8.13	80.4	
P3560-09	E1HY5	9	1.16	8.67	9.83	8.38	83.3	
P3560-10	E1HY6	10	1.19	8.71	9.9	8.34	82.1	
P3560-11	E1J17	11	1.19	8.58	9.77	7.81	77.2	
P3560-12	E1J18	12	1.15	8.40	9.55	8.11	82.9	
P3560-13	E1J19	13	1.19	8.47	9.66	7.93	79.6	
P3560-14	E1J20	14	1.18	8.49	9.67	7.75	77.4	
P3560-15	E1J21	15	1.16	8.71	9.87	7.85	76.8	
P3560-16	E1J22	16	1.15	8.84	9.99	8.22	80.0	
P3560-17	E1J22D	17	1.15	8.84	9.99	8.22	80.0	
P3560-18	E1J22S	18	1.15	8.84	9.99	8.22	80.0	
P3560-19	E1J23	19	1.15	8.62	9.77	7.69	75.9	
P3560-20	E1J24	20	1.15	8.84	9.99	8.8	86.5	
P3560-21	E1J25	21	1.15	8.57	9.72	8.02	80.2	
P3560-22	E1J26	22	1.12	8.46	9.58	8.74	90.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$