

LB131979


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40049



A= 0.0000e+000

B= 2.5144e-004 slope

C= -1.9990e-003 y-intercept

Rho= 0.9998822

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	abs
0.00	0.000	0.008	0.008	39	0.000	39					2
0.05	0.050					442	0				-4
0.20	0.200	0.192	-0.008	773	0.0 %	773					3
2.50	2.500	2.570	0.070	10229	0.0 %	10229					1
5.00	5.000	4.935	-0.065	19633	0.0 %	19633					1
7.50	7.500	7.427	-0.073	29546	0.0 %	29546					1
10.0	10.000	10.068	0.068	40049	0.0 %	40049					1

LB131979

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	39	-		0 SFAM01.1	PPB	8/13/2024 15:35	Std
	0.2 S01	773	-		0.2 SFAM01.1	PPB	8/13/2024 15:38	Std
	2.5 S02	10229	-		2.5 SFAM01.1	PPB	8/13/2024 15:40	Std
	5 S03	19633	-		5 SFAM01.1	PPB	8/13/2024 15:42	Std
	7.5 S04	29546	-		7.5 SFAM01.1	PPB	8/13/2024 15:44	Std
	10 S05	40049	-		10 SFAM01.1	PPB	8/13/2024 15:47	Std
ICV073	ICV073	15898	3.9955 -		SFAM01.1	PPB	8/13/2024 15:49	SMPL
ICB073	ICB073	-164	-0.0432 -		SFAM01.1	PPB	8/13/2024 15:52	SMPL
CCV040	CCV040	19620	4.9313 -		SFAM01.1	PPB	8/13/2024 15:54	SMPL
CCB040	CCB040	-171	-0.045 -		SFAM01.1	PPB	8/13/2024 15:56	SMPL
PB162710BL	PBW710	3	-0.0012 -		SFAM01.1	PPB	8/13/2024 15:58	SMPL
P3541-01	MCOB39	151	0.036 -		SFAM01.1	PPB	8/13/2024 16:01	SMPL
P3541-02	MCOB40	18	0.0025 -		SFAM01.1	PPB	8/13/2024 16:03	SMPL
P3541-03	MCOB42	18	0.0025 -		SFAM01.1	PPB	8/13/2024 16:05	SMPL
P3541-04	MCOB42D	29	0.0053 -		SFAM01.1	PPB	8/13/2024 16:08	SMPL
P3541-05	MCOB42S	4545	1.1408 -		SFAM01.1	PPB	8/13/2024 16:13	SMPL
P3541-06	MCOB44	-19	-0.0068 -		SFAM01.1	PPB	8/13/2024 16:15	SMPL
CCV041	CCV041	19628	4.9333 -		SFAM01.1	PPB	8/13/2024 16:17	SMPL
CCB041	CCB041	-162	-0.0427 -		SFAM01.1	PPB	8/13/2024 16:19	SMPL