

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131987

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/13/2024	14:30
2	CAL2	1	Water	NA	NA	20.3	7.01	08/13/2024	14:31
3	CAL3	1	Water	NA	NA	20.3	10.02	08/13/2024	14:33
4	ICV	1	Water	NA	NA	20.3	7.02	08/13/2024	14:35
5	CCV1	1	Water	NA	NA	20.2	2.01	08/13/2024	14:37
6	P3197-01	1	Solid	20.02	20	20.4	5.02	08/13/2024	14:38
7	P3577-01	1	Solid	20.02	20	20.4	6.33	08/13/2024	14:40
8	P3597-02	1	Solid	20.02	20	20.2	8.81	08/13/2024	14:45
9	P3597-02DUP	1	Solid	20.04	20	20.4	8.82	08/13/2024	14:46
10	CCV2	1	Water	NA	NA	20.3	12.02	08/13/2024	14:50

WORKLIST(Hardcopy Internal Chain)

131987

WorkList Name : corrosivity p3597 WorkList ID : 182614 Department : Wet-Chemistry Date : 08-13-2024 14:06:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3197-01	OCAP1002-01	Solid	Corrosivity	Cool 4 deg C	CHEM02	QA Of	07/10/2024	9045D
P3577-01	OILY-DEBRIS	Solid	Corrosivity	Cool 4 deg C	PSEG03	D21	08/12/2024	9045D
P3597-02	SP-1-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	D21	08/13/2024	9045D

Date/Time 08/13/24 14:20
Raw Sample Received by: JH WOOD
Raw Sample Relinquished by: JH WOOD

Date/Time 08/13/24 16:30
Raw Sample Received by: JH WOOD
Raw Sample Relinquished by: JH WOOD