

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131992

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	08/13/2024	09:30
2	P3574-02		1	11	100.00	>212.0	>212.0	08/13/2024	10:00
3	P3574-02DUP		1	11	100.00	>212.0	>212.0	08/13/2024	10:30
4	P3575-01		1	14	100.00	>212.0	>212.0	08/13/2024	11:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6131992

WorkList Name : fp-8-13 WorkList ID : 182582 Department : Wet-Chemistry Date : 08-13-2024 07:53:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3574-02	WC-W-06-202408	Water	Flash Point	Cool 4 deg C	AECO02	D21	08/12/2024	1010B
P3575-01	50350	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/12/2024	1010B

Date/Time 08/13/2024 09:20
Raw Sample Received by: RM LWCJ
Raw Sample Relinquished by: J Carey

Date/Time 08/13/2024 11:40
Raw Sample Received by: JWL
Raw Sample Relinquished by: RM LWCJ