

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131998

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	08/14/2024	12:10
2	P3576-05		1	11	100.00	>212.0	>212.0	08/14/2024	12:40
3	P3576-05DUP		1	11	100.00	>212.0	>212.0	08/14/2024	13:10
4	P3576-06		1	13	100.00	>212.0	>212.0	08/14/2024	13:40
5	P3577-02		1	6	9.00	48.2	47.5	08/14/2024	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6131998

WorkList Name : fp-8-14

WorkList ID : 182627

Department : Wet-Chemistry

Date : 08-14-2024 07:51:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3576-05	DW-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/12/2024	1010B
P3576-06	NWB-2008	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/12/2024	1010B
P3577-02	OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/12/2024	1010B

Date/Time 08/14/2024 12:00
Raw Sample Received by: RM(wc)
Raw Sample Relinquished by: RM(wc)

Date/Time 08/14/2024 11:45
Raw Sample Received by: RM(wc)
Raw Sample Relinquished by: RM(wc)