

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132020

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/15/2024	08:00
2	CAL2	1	Water	NA	NA	20.3	7.00	08/15/2024	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	08/15/2024	08:02
4	ICV	1	Water	NA	NA	20.3	7.01	08/15/2024	08:05
5	CCV1	1	Water	NA	NA	20.2	2.01	08/15/2024	08:10
6	P3611-06	1	Solid	20.02	20	21.4	7.16	08/15/2024	08:15
7	P3612-01	1	Solid	20.02	20	21.3	7.08	08/15/2024	08:20
8	P3612-01DUP	1	Solid	20.03	20	21.3	7.09	08/15/2024	08:21
9	CCV2	1	Water	NA	NA	20.3	12.02	08/15/2024	08:25

WORKLIST(Hardcopy Internal Chain)

UB 132020

WorkList Name : corrosivity p3611

WorkList ID : 182668

Department : Wet-Chemistry

Date : 08-15-2024 07:31:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3611-06	264	Solid	Corrosivity	Cool 4 deg C	PSEG03	D21	08/14/2024	9045D
P3612-01	409	Solid	Corrosivity	Cool 4 deg C	PSEG03	D21	08/14/2024	9045D

Date/Time 08/15/24 07:40

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 08/15/24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]