

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132022

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/15/2024	08:00
2	CAL2	1	Water	NA	NA	20.3	7.00	08/15/2024	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	08/15/2024	08:02
4	ICV	1	Water	NA	NA	20.3	7.01	08/15/2024	08:05
5	CCV1	1	Water	NA	NA	20.2	2.02	08/15/2024	08:30
6	P3197-02	1	Water	NA	NA	20.1	5.00	08/15/2024	08:33
7	P3577-02	1	Water	NA	NA	22.3	6.06	08/15/2024	08:35
8	P3591-01	1	Water	NA	NA	20.2	8.10	08/15/2024	08:40
9	P3591-02	1	Water	NA	NA	20.3	6.96	08/15/2024	08:45
10	P3610-01	1	Water	NA	NA	20.2	7.03	08/15/2024	08:50
11	P3610-02	1	Water	NA	NA	20.9	7.19	08/15/2024	09:00
12	P3611-01	1	Water	NA	NA	20.8	7.14	08/15/2024	09:10
13	P3611-02	1	Water	NA	NA	21.0	7.24	08/15/2024	09:15
14	P3611-07	1	Water	NA	NA	21.3	8.98	08/15/2024	09:25
15	P3611-07DUP	1	Water	NA	NA	21.4	8.99	08/15/2024	09:26
16	CCV2	1	Water	NA	NA	20.3	12.02	08/15/2024	09:30

WORKLIST(Hardcopy Internal Chain)

WB 132022

WorkList Name : ph p3611 WorkList ID : 182667 Department : Wet-Chemistry Date : 08-15-2024 07:30:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3197-02	OCAP1002-02	Water	pH	Cool 4 deg C	CHEM02	QA Of	07/10/2024	9040C
P3577-02	A OILY-WATER	Water	pH	Cool 4 deg C	PSEG03	D21	08/12/2024	9040C
P3591-01	PRODUCTION	Water	pH	Cool 4 deg C	SCIE01	D21	08/13/2024	9040C
P3591-02	PILOT-PLANT	Water	pH	Cool 4 deg C	SCIE01	D21	08/13/2024	9040C
P3610-01	A A6934	Water	pH	Cool 4 deg C	PSEG03	D31	08/14/2024	9040C
P3610-02	A M21005	Water	pH	Cool 4 deg C	PSEG03	D31	08/14/2024	9040C
P3611-01	A 258	Water	pH	Cool 4 deg C	PSEG03	D21	08/14/2024	9040C
P3611-02	A 259	Water	pH	Cool 4 deg C	PSEG03	D21	08/14/2024	9040C
P3611-07	A 255	Water	pH	Cool 4 deg C	PSEG03	D21	08/14/2024	9040C

Date/Time 08/15/24 07:45
 Raw Sample Received by: Shaw
 Raw Sample Relinquished by: Shaw

Date/Time 08/15/24 10:30
 Raw Sample Received by: Shaw
 Raw Sample Relinquished by: Shaw