

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132035

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	08/15/2024	11:20
2	P3611-03		1	4	8.00	46.4	46.1	08/15/2024	11:50
3	P3611-04		1	4	7.00	44.6	44.3	08/15/2024	12:30
4	P3611-05		1	4	7.00	44.6	44.3	08/15/2024	13:00
5	P3611-08		1	6	63.00	145.4	145.1	08/15/2024	13:30
6	P3611-08DUP		1	6	63.00	145.4	145.1	08/15/2024	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

U6132035

WorkList Name : FP-S8-14

WorkList ID : 182665

Department : Wet-Chemistry

Date : 08-14-2024 17:47:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3611-03 B	260	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	08/14/2024	1010B
P3611-04 A	261	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	08/14/2024	1010B
P3611-05 A	262	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	08/14/2024	1010B
P3611-08 A	253	Solid	Flash Point	Cool 4 deg C	PSEG03	D21	08/14/2024	1010B

Date/Time 08/15/2024 11.10
 Raw Sample Received by: RM
 Raw Sample Relinquished by: [Signature]

Date/Time 08/15/2024 14.46
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: RM (w.c.)