

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132043

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/15/2024	17:00
2	CAL2	1	Water	NA	NA	20.3	7.00	08/15/2024	17:01
3	CAL3	1	Water	NA	NA	20.4	10.02	08/15/2024	17:02
4	ICV	1	Water	NA	NA	20.2	7.01	08/15/2024	17:05
5	CCV1	1	Water	NA	NA	20.3	2.01	08/15/2024	17:38
6	P3622-01	1	Water	NA	NA	20.3	7.21	08/15/2024	17:40
7	P3622-01DUP	1	Water	NA	NA	20.4	7.22	08/15/2024	17:41
8	P3628-01	1	Water	NA	NA	21.3	7.38	08/15/2024	17:45
9	P3629-01	1	Water	NA	NA	21.2	5.76	08/15/2024	17:48
10	P3629-02	1	Water	NA	NA	21.3	5.69	08/15/2024	17:50
11	P3632-01	1	Water	NA	NA	22.3	7.24	08/15/2024	17:55
12	CCV2	1	Water	NA	NA	20.3	12.02	08/15/2024	18:00

WORKLIST(Hardcopy Internal Chain)

W 132043

WorkList Name : ph water p3628

WorkList ID : 182709

Department : Wet-Chemistry

Date : 08-15-2024 16:43:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3622-01	F 028-OUTFALL	Water	pH	Cool 4 deg C	MINO01	G21	08/13/2024	9040C
P3628-01	C 371	Water	pH	Cool 4 deg C	PSEG03	G22	08/15/2024	9040C
P3629-01	C 50405	Water	pH	Cool 4 deg C	PSEG03	G22	08/15/2024	9040C
P3629-02	C 50398	Water	pH	Cool 4 deg C	PSEG03	G22	08/15/2024	9040C
P3632-01	A 81224	Water	pH	Cool 4 deg C	PSEG03	G22	08/15/2024	9040C

Date/Time 08/15/24 16:50
 Raw Sample Received by: SO wep
 Raw Sample Relinquished by: Al sm

Date/Time 08/15/24 18:30
 Raw Sample Received by: Al sm
 Raw Sample Relinquished by: SO wep