

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132051

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	08/16/2024	10:30
2	P3610-01		1	12	100.00	>212.0	>212.0	08/16/2024	11:35
3	P3610-01DUP		1	12	100.00	>212.0	>212.0	08/16/2024	12:05
4	P3610-02		1	14	100.00	>212.0	>212.0	08/16/2024	12:35
5	P3611-01		1	11	100.00	>212.0	>212.0	08/16/2024	13:05
6	P3611-02		1	12	100.00	>212.0	>212.0	08/16/2024	13:35
7	P3611-07		1	14	100.00	>212.0	>212.0	08/16/2024	14:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26132051

WorkList Name : FP8-14

WorkList ID : 182664

Department : Wet-Chemistry

Date : 08-14-2024 17:47:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3610-01	A6934	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/14/2024	1010B
P3610-02	M21005	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/14/2024	1010B
P3611-01	258	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/14/2024	1010B
P3611-02	259	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/14/2024	1010B
P3611-07	255	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/14/2024	1010B

Date/Time 08/16/2024 10:25
 Raw Sample Received by: RM
 Raw Sample Relinquished by: 206001

Date/Time 08/16/2024 14:40
 Raw Sample Received by: 206001
 Raw Sample Relinquished by: RM (WC)