

Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132056

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/16/2024	16:40
2	CAL2	1	Water	NA	NA	20.2	7.01	08/16/2024	16:41
3	CAL3	1	Water	NA	NA	20.2	10.02	08/16/2024	16:44
4	ICV	1	Water	NA	NA	20.3	7.01	08/16/2024	16:45
5	CCV1	1	Water	NA	NA	20.3	2.01	08/16/2024	16:50
6	P3656-01	1	Water	NA	NA	20.2	6.21	08/16/2024	17:00
7	P3656-03	1	Water	NA	NA	20.9	6.71	08/16/2024	17:10
8	P3656-05	1	Water	NA	NA	20.9	6.72	08/16/2024	17:15
9	P3656-07	1	Water	NA	NA	20.7	6.78	08/16/2024	17:20
10	P3656-07DUP	1	Water	NA	NA	20.8	6.79	08/16/2024	17:21
11	CCV2	1	Water	NA	NA	20.3	12.02	08/16/2024	17:22

VB 132056

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p3656 WorkList ID : 182760 Department : Wet-Chemistry Date : 08-16-2024 16:28:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3656-01	MW-1	Water	pH	Cool 4 deg C	LOCK01	G11	08/16/2024	SM 4500-H B
P3656-03	MW-3	Water	pH	Cool 4 deg C	LOCK01	G11	08/16/2024	SM 4500-H B
P3656-05	MW-4	Water	pH	Cool 4 deg C	LOCK01	G11	08/16/2024	SM 4500-H B
P3656-07	MW-2	Water	pH	Cool 4 deg C	LOCK01	G11	08/16/2024	SM 4500-H B

Date/Time 08/16/24 16:35
Raw Sample Received by: SD Woy
Raw Sample Relinquished by: RM Cwu

Date/Time 08/16/24 18:00
Raw Sample Received by: RM Cwu
Raw Sample Relinquished by: SD Woy