

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132057

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/16/2024	16:40
2	CAL2	1	Water	NA	NA	20.2	7.01	08/16/2024	16:41
3	CAL3	1	Water	NA	NA	20.2	10.02	08/16/2024	16:44
4	ICV	1	Water	NA	NA	20.3	7.01	08/16/2024	16:45
5	CCV1	1	Water	NA	NA	20.4	2.02	08/16/2024	17:25
6	P3636-01	1	Solid	20.02	20	21.4	5.63	08/16/2024	17:30
7	P3637-02	1	Solid	20.03	20	20.3	6.83	08/16/2024	17:35
8	P3652-01	1	Solid	20.02	20	21.2	8.06	08/16/2024	17:38
9	P3652-01DUP	1	Solid	20.03	20	21.3	8.07	08/16/2024	17:39
10	CCV2	1	Water	NA	NA	20.3	12.02	08/16/2024	17:40

WORKLIST(Hardcopy Internal Chain)

132057

WorkList Name : corrosivity p3637 WorkList ID : 182761 Department : Wet-Chemistry Date : 08-16-2024 16:29:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3636-01	OILY-DEBRIS	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	08/16/2024	9045D
P3637-02	ETGI - 370	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	08/16/2024	9045D
P3652-01	LIVONIA-SOIL	Solid	Corrosivity	Cool 4 deg C	TULL02	G11	08/16/2024	9045D

Date/Time 08/16/24 16:35
Raw Sample Received by: JH CWC
Raw Sample Relinquished by: JH

Date/Time 08/16/24 18:30
Raw Sample Received by: JH
Raw Sample Relinquished by: JH