

LB132061


SFAM01.1

INSTRUMENT 1D1 CV1

Linear

 μ Abs.:

38513



A= 0.0000e+000

B= 2.6031e-004 slope

C= -8.9355e-002 y-intercept

Rho= 0.9999049

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.00	0.000	-0.019	-0.019	270	0.000	270					=
0.05	0.050					442	0				-12
0.20	0.200	0.177	-0.023	1025	0.0 %	1025					-1
2.50	2.500	2.483	-0.017	9883	0.0 %	9883					2
5.00	5.000	5.099	0.099	19930	0.0 %	19930					0
7.50	7.500	7.524	0.024	29246	0.0 %	29246					-1
10.0	10.000	9.936	-0.064	38513	0.0 %	38513					

LB132061
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	270	-		0 SFAM01.1	PPB	8/18/2024 12:42	Std
	0.2 S01	1025	-		0.2 SFAM01.1	PPB	8/18/2024 12:45	Std
	2.5 S02	9883	-		2.5 SFAM01.1	PPB	8/18/2024 12:47	Std
	5 S03	19930	-		5 SFAM01.1	PPB	8/18/2024 12:49	Std
	7.5 S04	29246	-		7.5 SFAM01.1	PPB	8/18/2024 12:52	Std
	10 S05	38513	-		10 SFAM01.1	PPB	8/18/2024 12:54	Std
ICV081	ICV081	15311	3.8963 -		SFAM01.1	PPB	8/18/2024 12:57	SMPL
ICB081	ICB081	19	-0.0844 -		SFAM01.1	PPB	8/18/2024 12:59	SMPL
CCV069	CCV069	19270	4.9268 -		SFAM01.1	PPB	8/18/2024 13:01	SMPL
CCB069	CCB069	10	-0.0868 -		SFAM01.1	PPB	8/18/2024 13:03	SMPL
PB162808BL	PBS808	19	-0.0844 -		SFAM01.1	PPB	8/18/2024 13:06	SMPL
P3423-01	MH1X86	5930	1.4543 -		SFAM01.1	PPB	8/18/2024 13:08	SMPL
P3423-02	MH1X88	14969	3.8072 -		SFAM01.1	PPB	8/18/2024 13:10	SMPL
P3423-03	MH1X90	8035	2.0022 -		SFAM01.1	PPB	8/18/2024 13:13	SMPL
P3423-04	MH1X92	10268	2.5835 -		SFAM01.1	PPB	8/18/2024 13:15	SMPL
P3423-05	MH1X94	850	0.1319 -		SFAM01.1	PPB	8/18/2024 13:17	SMPL
P3423-06	MH1X96	12109	3.0627 -		SFAM01.1	PPB	8/18/2024 13:19	SMPL
P3423-07	MH1X98	28912	7.4367 -		SFAM01.1	PPB	8/18/2024 13:22	SMPL
P3423-08	MH1Y01	4335	1.0391 -		SFAM01.1	PPB	8/18/2024 13:24	SMPL
P3423-09	MH1Y03	9255	2.3198 -		SFAM01.1	PPB	8/18/2024 13:26	SMPL
P3423-10	MH1Y05	6957	1.7216 -		SFAM01.1	PPB	8/18/2024 13:28	SMPL
P3423-11	MH1Y07	11138	2.81 -		SFAM01.1	PPB	8/18/2024 13:31	SMPL
P3423-12	MH1Y09	16187	4.1243 -		SFAM01.1	PPB	8/18/2024 13:33	SMPL
P3423-13	MH1Y11	3673	0.8668 -		SFAM01.1	PPB	8/18/2024 13:35	SMPL
P3423-14	MH1Y13	4895	1.1849 -		SFAM01.1	PPB	8/18/2024 13:38	SMPL
P3423-15	MH1Y15	22393	5.7398 -		SFAM01.1	PPB	8/18/2024 13:40	SMPL
P3423-16	MH1Y17	1859	0.3946 -		SFAM01.1	PPB	8/18/2024 13:42	SMPL
P3423-17	MH1Y19	10990	2.7715 -		SFAM01.1	PPB	8/18/2024 13:44	SMPL
P3423-18	MH1Y21	10472	2.6366 -		SFAM01.1	PPB	8/18/2024 13:47	SMPL
P3423-19	MH1Y23	1922	0.411 -		SFAM01.1	PPB	8/18/2024 13:49	SMPL
CCV070	CCV070	19866	5.082 -		SFAM01.1	PPB	8/18/2024 13:51	SMPL
CCB070	CCB070	86	-0.067 -		SFAM01.1	PPB	8/18/2024 13:53	SMPL
P3423-20	MH1Y23D	2077	0.4513 -		SFAM01.1	PPB	8/18/2024 13:56	SMPL
P3423-21	MH1Y23S	9849	2.4744 -		SFAM01.1	PPB	8/18/2024 13:58	SMPL
PB162809BL	PBS809	-3	-0.0901 -		SFAM01.1	PPB	8/18/2024 14:00	SMPL
P3425-01	MH1Y25	3127	0.7246 -		SFAM01.1	PPB	8/18/2024 14:03	SMPL
P3425-02	MH1Y27	2323	0.5153 -		SFAM01.1	PPB	8/18/2024 14:05	SMPL
P3425-03	MH1Y29	5090	1.2356 -		SFAM01.1	PPB	8/18/2024 14:07	SMPL
P3425-04	MH1Y31	3656	0.8623 -		SFAM01.1	PPB	8/18/2024 14:09	SMPL
P3425-05	MH1Y33	8416	2.1014 -		SFAM01.1	PPB	8/18/2024 14:12	SMPL
P3425-06	MH1Y35	1220	0.2282 -		SFAM01.1	PPB	8/18/2024 14:14	SMPL
P3425-07	MH1Y37	14608	3.7133 -		SFAM01.1	PPB	8/18/2024 14:16	SMPL
P3425-08	MH1Y39	6440	1.587 -		SFAM01.1	PPB	8/18/2024 14:18	SMPL
P3425-09	MH1Y41	1661	0.343 -		SFAM01.1	PPB	8/18/2024 14:21	SMPL
P3425-10	MH1Y41D	1696	0.3521 -		SFAM01.1	PPB	8/18/2024 14:31	SMPL
P3425-11	MH1Y41S	10199	2.5656 -		SFAM01.1	PPB	8/18/2024 14:33	SMPL

LB132061

INSTRUMENT ID : CV1

CCV071	CCV071	20176	5.1627 -	SFAM01.1	PPB	8/18/2024 14:35	SMPL
CCB071	CCB071	78	-0.0691 -	SFAM01.1	PPB	8/18/2024 14:38	SMPL