

LB132062

7470A

INSTRUMENT ID: CV3

Linear

 μ Abs.:

36369

A= 0.0000e+000

B= 2.7537e-004

C= -2.8606e-002

Rho= 0.9999943

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	-0.017	-0.017	43	0.000	43	0				15000
0.2	0.200	0.209	0.009	864	0.0 %	864					40000
2.5	2.500	2.507	0.007	9207	0.0 %	9207					00000
5.0	5.000	4.996	-0.004	18248	0.0 %	18248					
7.5	7.500	7.518	0.018	27405	0.0 %	27405					
10.0	10.000	9.986	-0.014	36369	0.0 %	36369					

LB132062 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Aug 2024 16:10:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.0000	7470A	PPB	18 Aug 2024 16:11:00	S	Std
0.2 - 1	50.2	864	-	0.2000	7470A	PPB	18 Aug 2024 16:13:16	S	Std
2.5 - 1	52.5	9207	-	2.5000	7470A	PPB	18 Aug 2024 16:15:33	S	Std
5.0 - 1	55	18248	-	5.0000	7470A	PPB	18 Aug 2024 16:17:50	S	Std
7.5 - 1	57.5	27405	-	7.5000	7470A	PPB	18 Aug 2024 16:20:06	S	Std
10.0 - 1	510	36369	-	10.0000	7470A	PPB	18 Aug 2024 16:27:40	S	Std
ICV82 - 1	ICV82	13964	3.8167	-	7470A	PPB	18 Aug 2024 16:30:33	U	SMPL
ICB82 - 1	ICB82	-112	-0.0594	-	7470A	PPB	18 Aug 2024 16:32:47	U	SMPL
CCV72 - 1	CCV72	18235	4.9928	-	7470A	PPB	18 Aug 2024 16:35:04	U	SMPL
CCB72 - 1	CCB72	-175	-0.0768	-	7470A	PPB	18 Aug 2024 16:37:18	U	SMPL
CRA - 1	CRA	826	0.1989	-	7470A	PPB	18 Aug 2024 16:39:34	U	SMPL
HighStd - 1	HighStd	35223	9.6708	-	7470A	PPB	18 Aug 2024 16:41:48	U	SMPL
ChkStd - 1	ChkStd	26074	7.1514	-	7470A	PPB	18 Aug 2024 16:44:03	U	SMPL
PB162807BL - 1	PBW	-232	-0.0925	-	7470A	PPB	18 Aug 2024 16:46:20	U	SMPL
PB162807BS - 1	LCSW	16155	4.4200	-	7470A	PPB	18 Aug 2024 16:48:38	U	SMPL
P3611-01 - 1	258	-117	-0.0608	-	7470A	PPB	18 Aug 2024 16:50:54	U	SMPL
P3611-01DUP - 1	258DUP	44	-0.0165	-	7470A	PPB	18 Aug 2024 16:53:11	U	SMPL
P3611-01MS - 1	258MS	13792	3.7693	-	7470A	PPB	18 Aug 2024 16:55:28	U	SMPL
P3611-01MSD - 1	258MSD	14493	3.9624	-	7470A	PPB	18 Aug 2024 16:57:44	U	SMPL
P3611-02 - 1	259	-136	-0.0661	-	7470A	PPB	18 Aug 2024 17:00:00	U	SMPL
CCV73 - 1	CCV73	18440	5.0492	-	7470A	PPB	18 Aug 2024 17:02:17	U	SMPL
CCB73 - 1	CCB73	-154	-0.0710	-	7470A	PPB	18 Aug 2024 17:04:32	U	SMPL
P3611-07 - 1	255	59	-0.0124	-	7470A	PPB	18 Aug 2024 17:06:49	U	SMPL
PB162810BL - 1	PBW	41	-0.0173	-	7470A	PPB	18 Aug 2024 17:09:04	U	SMPL
PB162810BS - 1	LCSW	12841	3.5074	-	7470A	PPB	18 Aug 2024 17:11:19	U	SMPL
P3611-03 - 1	260	1462	0.3740	-	7470A	PPB	18 Aug 2024 17:13:35	U	SMPL
P3611-04 - 1	261	160	0.0155	-	7470A	PPB	18 Aug 2024 17:15:51	U	SMPL
P3611-05 - 1	262	1079	0.2685	-	7470A	PPB	18 Aug 2024 17:18:09	U	SMPL
P3611-06 - 1	264	-180	-0.0782	-	7470A	PPB	18 Aug 2024 17:20:25	U	SMPL
P3611-08 - 1	253	218	0.0314	-	7470A	PPB	18 Aug 2024 17:22:43	U	SMPL
P3612-01 - 1	409	-326	-0.1184	-	7470A	PPB	18 Aug 2024 17:24:59	U	SMPL
P3613-01 - 1	TRE-1802	-82	-0.0512	-	7470A	PPB	18 Aug 2024 17:27:16	U	SMPL
CCV74 - 1	CCV74	17918	4.9055	-	7470A	PPB	18 Aug 2024 17:29:31	U	SMPL
CCB74 - 1	CCB74	-232	-0.0925	-	7470A	PPB	18 Aug 2024 17:31:47	U	SMPL
P3629-03 - 1	SB-1	324	0.0606	-	7470A	PPB	18 Aug 2024 17:34:03	U	SMPL
P3629-04 - 1	SB-2	-80	-0.0506	-	7470A	PPB	18 Aug 2024 17:36:19	U	SMPL
P3629-05 - 1	SB-3	-48	-0.0418	-	7470A	PPB	18 Aug 2024 17:38:34	U	SMPL
P3629-06 - 1	SB-4	-25	-0.0355	-	7470A	PPB	18 Aug 2024 17:40:50	U	SMPL
P3629-07 - 1	SB-5	329	0.0620	-	7470A	PPB	18 Aug 2024 17:43:05	U	SMPL
P3631-01 - 1	358	-113	-0.0597	-	7470A	PPB	18 Aug 2024 17:45:21	U	SMPL
P3640-02 - 1	SP-1-COMP	0	-0.0286	-	7470A	PPB	18 Aug 2024 17:47:38	U	SMPL
P3640-02DUP - 1	SP-1-COMPDUP	1	-0.0283	-	7470A	PPB	18 Aug 2024 17:49:54	U	SMPL
P3640-02MS - 1	SP-1-COMPMS	12901	3.5240	-	7470A	PPB	18 Aug 2024 17:52:11	U	SMPL
P3640-02MSD - 1	SP-1-COMPMSD	13437	3.6716	-	7470A	PPB	18 Aug 2024 17:54:26	U	SMPL
CCV75 - 1	CCV75	17710	4.8482	-	7470A	PPB	18 Aug 2024 17:56:42	U	SMPL
CCB75 - 1	CCB75	-186	-0.0798	-	7470A	PPB	18 Aug 2024 17:58:59	U	SMPL
PB162748TB - 1	PB162748TB	-31	-0.0371	-	7470A	PPB	18 Aug 2024 18:01:16	U	SMPL
PB162749TB - 1	PB162749TB	25	-0.0217	-	7470A	PPB	18 Aug 2024 18:03:32	U	SMPL
P3611-01LX5 - 1		5	-0.0272	-	7470A	PPB	18 Aug 2024 18:05:47	U	SMPL
P3611-01A - 1		15510	4.2424	-	7470A	PPB	18 Aug 2024 18:08:02	U	SMPL
P3640-02LX5 - 1		-133	-0.0652	-	7470A	PPB	18 Aug 2024 18:10:18	U	SMPL
P3640-02A - 1		15354	4.1994	-	7470A	PPB	18 Aug 2024 18:12:35	U	SMPL
CCV76 - 1	CCV76	18160	4.9721	-	7470A	PPB	18 Aug 2024 18:14:51	U	SMPL
CCB76 - 1	CCB76	-206	-0.0853	-	7470A	PPB	18 Aug 2024 18:17:09	U	SMPL