

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132068

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/19/2024	13:15
2	P3628-01		1	14	100.00	>212.0	>212.0	08/19/2024	13:45
3	P3629-01		1	12	100.00	>212.0	>212.0	08/19/2024	14:45
4	P3629-02		1	11	100.00	>212.0	>212.0	08/19/2024	15:15
5	P3632-01		1	12	100.00	>212.0	>212.0	08/19/2024	16:00
6	P3632-01DUP		1	12	100.00	>212.0	>212.0	08/19/2024	16:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132068

WorkList Name : FP-8-16

WorkList ID : 182755

Department : Wet-Chemistry

Date : 08-16-2024 14:12:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3628-01	371	Water	Flash Point	Cool 4 deg C	PSEG03	G22	08/15/2024	1010B
P3629-01	50405	Water	Flash Point	Cool 4 deg C	PSEG03	G22	08/15/2024	1010B
P3629-02	50398	Water	Flash Point	Cool 4 deg C	PSEG03	G22	08/15/2024	1010B
P3632-01	81224	Water	Flash Point	Cool 4 deg C	PSEG03	G22	08/15/2024	1010B

Date/Time 08/19/2024 13:10
 Raw Sample Received by: RM (w/c)
 Raw Sample Relinquished by: RM (w/c)

Date/Time 08/19/2024 17:05
 Raw Sample Received by: RM (w/c)
 Raw Sample Relinquished by: RM (w/c)