

LB13209B

245.1

INSTRUMENT 1A: CV1

Linear

 μ Abs.:

37697

A= 0.0000e+000

B= 2.6709e-004

C= -7.5551e-003

Rho= 0.9998897

Accept=Accepted

Max. Conc.: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/v D
0.0	0.000	0.003	0.003	39	0.000	39					0/0 D
0.05	0.050	0.061	0.011	256	0.0 %	256	0				21
0.2	0.200	0.201	0.001	779	0.0 %	779					1
2.5	2.500	2.485	-0.015	9331	0.0 %	9331					1
5.0	5.000	5.058	0.058	18965	0.0 %	18965					1
7.5	7.500	7.382	-0.118	27668	0.0 %	27668					1
10.0	10.000	10.061	0.061	37697	0.0 %	37697					21

LB132093 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Aug 2024 13:45:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	245.1	PPB	20 Aug 2024 13:53:50	S	Std
0.05 - 1	50.05	256	-	0.0500	245.1	PPB	20 Aug 2024 13:56:06	S	Std
0.2 - 1	50.2	779	-	0.2000	245.1	PPB	20 Aug 2024 13:58:23	S	Std
2.5 - 1	52.5	9331	-	2.5000	245.1	PPB	20 Aug 2024 14:00:39	S	Std
5.0 - 1	55	18965	-	5.0000	245.1	PPB	20 Aug 2024 14:02:55	S	Std
7.5 - 1	57.5	27668	-	7.5000	245.1	PPB	20 Aug 2024 14:05:12	S	Std
10.0 - 1	510	37697	-	10.0000	245.1	PPB	20 Aug 2024 14:13:05	S	Std
ICV86 - 1	ICV86	14691	3.9163	-245.1		PPB	20 Aug 2024 14:18:11	U	SMPL
ICB86 - 1	ICB86	-153	-0.0484	-245.1		PPB	20 Aug 2024 14:20:25	U	SMPL
CCV87 - 1	CCV87	18971	5.0594	-245.1		PPB	20 Aug 2024 14:22:42	U	SMPL
CCB87 - 1	CCB87	-143	-0.0457	-245.1		PPB	20 Aug 2024 14:24:57	U	SMPL
CRA - 1	CRA	233	0.0547	-245.1		PPB	20 Aug 2024 14:27:13	U	SMPL
HighStd - 1	HighStd	38369	10.2405	-245.1		PPB	20 Aug 2024 14:29:29	U	SMPL
ChkStd - 1	ChkStd	24471	6.5284	-245.1		PPB	20 Aug 2024 14:31:44	U	SMPL
PB162865BL - 1	PBW	-190	-0.0583	-245.1		PPB	20 Aug 2024 14:34:02	U	SMPL
PB162865BS - 1	LCSW	15817	4.2170	-245.1		PPB	20 Aug 2024 14:36:20	U	SMPL
P3204-05 - 1	PT-HG-WS	26396	7.0426	-245.1		PPB	20 Aug 2024 14:38:36	U	SMPL
P3609-01 - 1	915-J-WS-081424	249	0.0590	-245.1		PPB	20 Aug 2024 14:40:54	U	SMPL
P3609-01DUP - 1	915-J-WS-081424DUP	79	0.0135	-245.1		PPB	20 Aug 2024 14:45:52	U	SMPL
P3609-01MS - 1	915-J-WS-081424MS	17264	4.6035	-245.1		PPB	20 Aug 2024 14:48:09	U	SMPL
P3609-01MSD - 1	915-J-WS-081424MSD	16409	4.3751	-245.1		PPB	20 Aug 2024 14:50:24	U	SMPL
CCV88 - 1	CCV88	19374	5.1671	-245.1		PPB	20 Aug 2024 14:52:40	U	SMPL
CCB88 - 1	CCB88	-137	-0.0441	-245.1		PPB	20 Aug 2024 14:54:57	U	SMPL
P3609-02 - 1	920-J-WS-081424	193	0.0440	-245.1		PPB	20 Aug 2024 14:57:15	U	SMPL
P3609-01LX5 - 1		1	-0.0073	-245.1		PPB	20 Aug 2024 14:59:30	U	SMPL
P3609-01A - 1		14548	3.8781	-245.1		PPB	20 Aug 2024 15:01:45	U	SMPL
CCV89 - 1	CCV89	19604	5.2285	-245.1		PPB	20 Aug 2024 15:08:36	U	SMPL
CCB89 - 1	CCB89	-182	-0.0562	-245.1		PPB	20 Aug 2024 15:10:56	U	SMPL