

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132094

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	08/20/2024	14:15
2	P3668-01		1	11	100.00	>212.0	>212.0	08/20/2024	15:20
3	P3668-01DUP		1	11	100.00	>212.0	>212.0	08/20/2024	15:50
4	P3669-03		1	12	100.00	>212.0	>212.0	08/20/2024	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6132094

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-8-20

WorkList ID : 182815

Department : Wet-Chemistry

Date : 08-20-2024 07:55:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3668-01	60349-60350	Water	Flash Point	Cool 4 deg C	PSEG03	G11	08/19/2024	1010B
P3669-03	1027	Water	Flash Point	Cool 4 deg C	PSEG03	G11	08/19/2024	1010B

Date/Time 08/20/2024 14.10
Raw Sample Received by: RM awc
Raw Sample Relinquished by: Mose

Date/Time 08/20/2024 17.00
Raw Sample Received by: RM awc
Raw Sample Relinquished by: RM awc