

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132100

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/21/2024	08:40
2	CAL2	1	Water	NA	NA	20.2	7.00	08/21/2024	08:41
3	CAL3	1	Water	NA	NA	20.3	10.02	08/21/2024	08:42
4	ICV	1	Water	NA	NA	20.3	7.01	08/21/2024	08:45
5	CCV1	1	Water	NA	NA	20.3	2.01	08/21/2024	09:35
6	P3668-01	1	Water	NA	NA	20.6	5.60	08/21/2024	09:40
7	P3668-01DUP	1	Water	NA	NA	20.7	5.61	08/21/2024	09:41
8	P3669-03	1	Water	NA	NA	20.1	6.77	08/21/2024	09:45
9	P3684-08	1	Water	NA	NA	20.3	7.13	08/21/2024	09:50
10	P3684-09	1	Water	NA	NA	20.1	8.35	08/21/2024	10:00
11	P3684-10	1	Water	NA	NA	20.3	6.90	08/21/2024	10:10
12	CCV2	1	Water	NA	NA	20.3	12.02	08/21/2024	10:15

WORKLIST(Hardcopy Internal Chain)

W3132100

WorkList Name : ph p3668

WorkList ID : 182839

Department : Wet-Chemistry

Date : 08-21-2024 08:23:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3668-01 A	60349-60350	Water	pH	Cool 4 deg C	PSEG03	G11	08/19/2024	9040C
P3668-03 C	1027	Water	pH	Cool 4 deg C	PSEG03	G11	08/19/2024	9040C
P3684-08 A	T3650	Water	pH	Cool 4 deg C	PSEG03	G11	08/20/2024	9040C
P3684-09 A	D3643	Water	pH	Cool 4 deg C	PSEG03	G11	08/20/2024	9040C
P3684-10 A	D3637	Water	pH	Cool 4 deg C	PSEG03	G11	08/20/2024	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/21/24 08:25

JP Sr

JP Sr

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/21/24

JP Sr

JP Sr

10:30

JP Sr

JP Sr