


CB132103

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.3659e-004 slope

C= -1.6872e-002 $y = 1.712e-005$

Rho= 0.9997290

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/e)
0.00	0.000	0.053	0.053	294	0.000	294					
0.05	0.050					442	0				
0.20	0.200	0.211	0.011	962	0.0 %	962					
2.50	2.500	2.537	0.037	10794	0.0 %	10794					
5.00	5.000	4.877	-0.123	20684	0.0 %	20684					
7.50	7.500	7.401	-0.099	31351	0.0 %	31351					
10.0	10.000	10.122	0.122	42855	0.0 %	42855					

o/e)
1
6
1
-2
-1
1

LB132103

INSTRUMENT ID : CV

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	294	-		0 SFAM01.1	PPB	8/21/2024 11:16	Std
	0.2 S01	962	-		0.2 SFAM01.1	PPB	8/21/2024 11:19	Std
	2.5 S02	10794	-		2.5 SFAM01.1	PPB	8/21/2024 11:21	Std
	5 S03	20684	-		5 SFAM01.1	PPB	8/21/2024 11:23	Std
	7.5 S04	31351	-		7.5 SFAM01.1	PPB	8/21/2024 11:25	Std
	10 S05	42855	-		10 SFAM01.1	PPB	8/21/2024 11:28	Std
ICV087	ICV087	16383	3.8592 -		SFAM01.1	PPB	8/21/2024 11:30	SMPL
ICB087	ICB087	65	-0.0015 -		SFAM01.1	PPB	8/21/2024 11:33	SMPL
CCV090	CCV090	20068	4.7311 -		SFAM01.1	PPB	8/21/2024 11:35	SMPL
CCB090	CCB090	52	-0.0046 -		SFAM01.1	PPB	8/21/2024 11:37	SMPL
PB162886BL	PBW886	48	-0.0055 -		SFAM01.1	PPB	8/21/2024 11:40	SMPL
PB162795TB	LEB795	88	0.0039 -		SFAM01.1	PPB	8/21/2024 11:42	SMPL
P3594-01	MDCXW2	872	0.1894 -		SFAM01.1	PPB	8/21/2024 11:44	SMPL
P3594-02	MDCXW3	1732	0.3929 -		SFAM01.1	PPB	8/21/2024 11:46	SMPL
P3594-03	MDCXW3D	1724	0.391 -		SFAM01.1	PPB	8/21/2024 11:49	SMPL
P3594-04	MDCXW3S	5145	1.2004 -		SFAM01.1	PPB	8/21/2024 11:56	SMPL
CCV091	CCV091	19288	4.5465 -		SFAM01.1	PPB	8/21/2024 11:59	SMPL
CCB091	CCB091	36	-0.0084 -		SFAM01.1	PPB	8/21/2024 12:01	SMPL