

LB132104

7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44894

A= 0.0000e+000

B= 2.2456e-004

C= -4.1994e-003

Rho= 0.9998467

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10
0.0	0.000	0.011	0.011	68	0.000	68					
0.2	0.200	0.197	-0.003	896	0.0 %	896					-2
2.5	2.500	2.515	0.015	11220	0.0 %	11220					1
5.0	5.000	5.032	0.032	22427	0.0 %	22427					1
7.5	7.500	7.367	-0.133	32825	0.0 %	32825					-2
10.0	10.000	10.077	0.077	44894	0.0 %	44894					1

LB132104 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Aug 2024 12:37:16

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	68	-	0.00007471B	PPB		21 Aug 2024 13:18:32	S	Std
0.2 - 1	50-2	896	-	0.20007471B	PPB		21 Aug 2024 13:20:46	S	Std
2.5 - 1	52-5	11220	-	2.50007471B	PPB		21 Aug 2024 13:23:06	S	Std
5.0 - 1	55	22427	-	5.00007471B	PPB		21 Aug 2024 13:25:22	S	Std
7.5 - 1	57-5	32825	-	7.50007471B	PPB		21 Aug 2024 13:27:41	S	Std
10.0 - 1	514	44894	-	10.00007471B	PPB		21 Aug 2024 13:33:02	S	Std
ICV88 - 1	ICV88	16538	3.7096	-7471B	PPB		21 Aug 2024 13:35:47	U	SMPL
ICB88 - 1	ICB88	-212	-0.0518	-7471B	PPB		21 Aug 2024 13:38:04	U	SMPL
CCV92 - 1	CCV92	22807	5.1174	-7471B	PPB		21 Aug 2024 13:40:21	U	SMPL
CCB92 - 1	CCB92	-233	-0.0565	-7471B	PPB		21 Aug 2024 13:42:36	U	SMPL
CRA - 1	CRA	885	0.1945	-7471B	PPB		21 Aug 2024 13:44:54	U	SMPL
HighStd - 1	HighStd	45817	10.2846	-7471B	PPB		21 Aug 2024 13:47:08	U	SMPL
ChkStd - 1	ChkStd	29032	6.5153	-7471B	PPB		21 Aug 2024 13:49:24	U	SMPL
PB162883BL - 1	PBS	-312	-0.0743	-7471B	PPB		21 Aug 2024 13:51:48	U	SMPL
PB162883BS - 1	LCSS	17971	4.0314	-7471B	PPB		21 Aug 2024 13:54:08	U	SMPL
P3641-01 - 1	S-907-K1-0-0.5-81524	6603	1.4786	-7471B	PPB		21 Aug 2024 13:56:25	U	SMPL
P3641-02 - 1	S-907-K1-0.5-1.0-81524	4978	1.1137	-7471B	PPB		21 Aug 2024 13:58:44	U	SMPL
P3641-03 - 1	S-908-K1-0-0.5-81524	4836	1.0818	-7471B	PPB		21 Aug 2024 14:01:03	U	SMPL
P3641-04 - 1	S-908-K1-0.5-1.0-81524	2748	0.6129	-7471B	PPB		21 Aug 2024 14:03:21	U	SMPL
P3641-05 - 1	S-909-K1-0-0.5-81524	3361	0.7506	-7471B	PPB		21 Aug 2024 14:05:37	U	SMPL
CCV93 - 1	CCV93	22161	4.9723	-7471B	PPB		21 Aug 2024 14:07:54	U	SMPL
CCB93 - 1	CCB93	-183	-0.0453	-7471B	PPB		21 Aug 2024 14:10:10	U	SMPL
P3641-06 - 1	S-909-K1-0.5-1.0-81524	3053	0.6814	-7471B	PPB		21 Aug 2024 14:12:30	U	SMPL
P3641-06MS - 1	S-909-K1-0.5-1.0-81524MS	19371	4.3458	-7471B	PPB		21 Aug 2024 14:14:45	U	SMPL
P3641-06MSD - 1	S-909-K1-0.5-1.0-81524MSD	19032	4.2697	-7471B	PPB		21 Aug 2024 14:17:03	U	SMPL
P3641-07 - 1	S-910-K1-0-0.5-81524	3338	0.7454	-7471B	PPB		21 Aug 2024 14:19:22	U	SMPL
P3641-08 - 1	S-910-K1-0.5-1.0-81524-FD	4223	0.9441	-7471B	PPB		21 Aug 2024 14:21:41	U	SMPL
P3641-09 - 1	S-910-K1-0.5-1.0-81524	3197	0.7137	-7471B	PPB		21 Aug 2024 14:24:02	U	SMPL
P3641-10 - 1	S-910-K1-0.5-1.0-81524-FD	3338	0.7454	-7471B	PPB		21 Aug 2024 14:26:20	U	SMPL
P3669-01 - 1	20430	758	0.1660	-7471B	PPB		21 Aug 2024 14:28:38	U	SMPL
P3685-01 - 1	Stock Pile-1	2107	0.4690	-7471B	PPB		21 Aug 2024 14:30:55	U	SMPL
P3685-03 - 1	Stock Pile-2	2020	0.4494	-7471B	PPB		21 Aug 2024 14:33:11	U	SMPL
CCV94 - 1	CCV94	22521	5.0532	-7471B	PPB		21 Aug 2024 14:35:26	U	SMPL
CCB94 - 1	CCB94	-200	-0.0491	-7471B	PPB		21 Aug 2024 14:37:42	U	SMPL
PB162884BL - 1	PB162884BL	-18	-0.0082	-7471B	PPB		21 Aug 2024 14:40:00	U	SMPL
PB162884BS - 1	PB162884BS	16469	3.6941	-7471B	PPB		21 Aug 2024 14:42:16	U	SMPL
P3643-01 - 1	932-K1-SD-0-0.5-081524	2953	0.6589	-7471B	PPB		21 Aug 2024 14:45:09	U	SMPL
P3643-01DUP - 1	932-K1-SD-0-0.5-081524DUP	2832	0.6318	-7471B	PPB		21 Aug 2024 14:47:34	U	SMPL
P3643-02 - 1	P3643-01MS	17521	3.9304	-7471B	PPB		21 Aug 2024 14:50:00	U	SMPL
P3643-03 - 1	P3643-01MSD	17497	3.9250	-7471B	PPB		21 Aug 2024 14:52:26	U	SMPL
P3643-04 - 1	932-K1-SD-0.5-1.0-081524	1055	0.2327	-7471B	PPB		21 Aug 2024 14:54:52	U	SMPL
P3643-05 - 1	917-J-SD-0-0.5-081524	19717	4.4235	-7471B	PPB		21 Aug 2024 14:57:17	U	SMPL
P3643-06 - 1	917-J-SD-0.5-1.0-081524	14233	3.1920	-7471B	PPB		21 Aug 2024 15:00:42	U	SMPL
P3643-07 - 1	918-J-SD-0-0.5-081524	7294	1.6338	-7471B	PPB		21 Aug 2024 15:03:07	U	SMPL
CCV95 - 1	CCV95	22401	5.0262	-7471B	PPB		21 Aug 2024 15:05:32	U	SMPL
CCB95 - 1	CCB95	-180	-0.0446	-7471B	PPB		21 Aug 2024 15:07:57	U	SMPL
P3643-08 - 1	918-J-SD-0.5-1.0-081524	8541	1.9138	-7471B	PPB		21 Aug 2024 15:10:22	U	SMPL
P3643-09 - 1	924-K1-SD-0-0.5-081524	1270	0.2810	-7471B	PPB		21 Aug 2024 15:12:47	U	SMPL
P3643-10 - 1	924-K1-SD-0.5-1.0-081524	427	0.0917	-7471B	PPB		21 Aug 2024 15:15:02	U	SMPL
P3643-11 - 1	921-J-SD-0-0.5-081524	19581	4.3930	-7471B	PPB		21 Aug 2024 15:17:27	U	SMPL
P3643-12 - 1	921-J-SD-0.5-1.0-081524	3194	0.7131	-7471B	PPB		21 Aug 2024 15:19:52	U	SMPL
P3643-13 - 1	921-J-SD-0-0.5-081524-FD	18302	4.1058	-7471B	PPB		21 Aug 2024 15:22:17	U	SMPL
P3643-14 - 1	921-K1-SD-0.5-1.0-081524-FD	24659	5.5333	-7471B	PPB		21 Aug 2024 15:24:42	U	SMPL
CCV96 - 1	CCV96	22051	4.9476	-7471B	PPB		21 Aug 2024 15:27:07	U	SMPL
CCB96 - 1	CCB96	-177	-0.0439	-7471B	PPB		21 Aug 2024 15:29:32	U	SMPL
P3641-06DUP - 1	S-907-K1-0-0.5-81524DUP	2820	0.6291	-7471B	PPB		21 Aug 2024 15:31:57	U	SMPL
P3641-06LX5 - 1		423	0.0908	-7471B	PPB		21 Aug 2024 15:34:22	U	SMPL
P3641-06A - 1		19233	4.3148	-7471B	PPB		21 Aug 2024 15:36:47	U	SMPL
P3643-01LX5 - 1		652	0.1422	-7471B	PPB		21 Aug 2024 15:39:12	U	SMPL
P3643-01A - 1		23040	5.1697	-7471B	PPB		21 Aug 2024 15:41:37	U	SMPL
CCV97 - 1	CCV97	21622	4.8513	-7471B	PPB		21 Aug 2024 15:43:62	U	SMPL
CCB97 - 1	CCB97	-176	-0.0437	-7471B	PPB		21 Aug 2024 15:45:87	U	SMPL