


LB132130

7470A

INSTRUMENT DISCU



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/21)
0.0	0.000	0.033	0.033	34	0.000	34					
0.2	0.200	0.247	0.047	901	0.0 %	901	0				24
2.5	2.500	2.451	-0.049	9837	0.0 %	9837					-2
5.0	5.000	4.776	-0.224	19262	0.0 %	19262					-4
7.5	7.500	7.869	0.369	31800	0.0 %	31800					5
10.0	10.000	9.824	-0.176	39724	0.0 %	39724					-2

LB132130 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Aug 2024 13:46:23

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7470A	PPB	22 Aug 2024 14:02:47	S	Std
0.2 - 1	50-2	901	-	0.2000	7470A	PPB	22 Aug 2024 14:05:04	S	Std
2.5 - 1	52.5	9837	-	2.5000	7470A	PPB	22 Aug 2024 14:07:20	S	Std
5.0 - 1	55	19262	-	5.0000	7470A	PPB	22 Aug 2024 14:09:37	S	Std
7.5 - 1	57.5	31800	-	7.5000	7470A	PPB	22 Aug 2024 14:11:54	S	Std
10.0 - 1	510	39724	-	10.0000	7470A	PPB	22 Aug 2024 14:22:27	S	Std
ICV92 - 1	ICV92	15240	3.7840	-	7470A	PPB	22 Aug 2024 14:26:17	U	SMPL
ICB92 - 1	ICB92	-99	-0.0000	-	7470A	PPB	22 Aug 2024 14:28:32	U	SMPL
CCV11 - 1	CCV11	19985	4.9545	-	7470A	PPB	22 Aug 2024 14:30:48	U	SMPL
CCB11 - 1	CCB11	-131	-0.0079	-	7470A	PPB	22 Aug 2024 14:33:03	U	SMPL
CRA - 1	CRA	848	0.2336	-	7470A	PPB	22 Aug 2024 14:35:19	U	SMPL
HighStd - 1	HighStd	39850	9.8551	-	7470A	PPB	22 Aug 2024 14:37:34	U	SMPL
ChkStd - 1	ChkStd	29336	7.2614	-	7470A	PPB	22 Aug 2024 14:39:49	U	SMPL
PB162892BL - 1	PBW	-237	-0.0340	-	7470A	PPB	22 Aug 2024 14:42:07	U	SMPL
PB162892BS - 1	LCSW	17519	4.3462	-	7470A	PPB	22 Aug 2024 14:44:25	U	SMPL
P3669-05 - 1	20430	-183	-0.0207	-	7470A	PPB	22 Aug 2024 14:46:42	U	SMPL
P3684-03 - 1	D3646	-66	0.0081	-	7470A	PPB	22 Aug 2024 14:48:59	U	SMPL
P3684-04 - 1	D3638	-8	0.0224	-	7470A	PPB	22 Aug 2024 14:51:16	U	SMPL
P3684-06 - 1	D3564	366	0.1147	-	7470A	PPB	22 Aug 2024 14:53:32	U	SMPL
P3685-02 - 1	Stock Pile-1	-78	0.0052	-	7470A	PPB	22 Aug 2024 14:55:48	U	SMPL
CCV12 - 1	CCV12	19465	4.8263	-	7470A	PPB	22 Aug 2024 14:58:02	U	SMPL
CCB12 - 1	CCB12	-142	-0.0106	-	7470A	PPB	22 Aug 2024 15:00:18	U	SMPL
P3685-04 - 1	Stock Pile-2	-16	0.0205	-	7470A	PPB	22 Aug 2024 15:02:35	U	SMPL
P3685-04DUP - 1	Stock Pile-2DUP	0	0.0244	-	7470A	PPB	22 Aug 2024 15:04:50	U	SMPL
P3685-04MS - 1	Stock Pile-2MS	14573	3.6195	-	7470A	PPB	22 Aug 2024 15:07:05	U	SMPL
P3685-04MSD - 1	Stock Pile-2MSD	14335	3.5607	-	7470A	PPB	22 Aug 2024 15:09:20	U	SMPL
PB162896BL - 1	PBW	-173	-0.0183	-	7470A	PPB	22 Aug 2024 15:11:38	U	SMPL
P3684-08 - 1	T3650	-58	0.0101	-	7470A	PPB	22 Aug 2024 15:16:11	U	SMPL
P3684-09 - 1	D3643	11	0.0271	-	7470A	PPB	22 Aug 2024 15:18:28	U	SMPL
P3684-10 - 1	D3637	0	0.0244	-	7470A	PPB	22 Aug 2024 15:20:45	U	SMPL
P3684-10DUP - 1	D3637DUP	12	0.0274	-	7470A	PPB	22 Aug 2024 15:23:00	U	SMPL
CCV13 - 1	CCV13	20353	5.0453	-	7470A	PPB	22 Aug 2024 15:25:15	U	SMPL
CCB13 - 1	CCB13	-144	-0.0111	-	7470A	PPB	22 Aug 2024 15:27:31	U	SMPL
P3684-10MS - 1	D3637MS	15656	3.8866	-	7470A	PPB	22 Aug 2024 15:35:25	U	SMPL
P3684-10MSD - 1	D3637MSD	15617	3.8770	-	7470A	PPB	22 Aug 2024 15:37:40	U	SMPL
PB162924BL - 1	PBW	-177	-0.0192	-	7470A	PPB	22 Aug 2024 15:39:57	U	SMPL
PB162924BS - 1	LCSW	14978	3.7194	-	7470A	PPB	22 Aug 2024 15:42:13	U	SMPL
P3669-03 - 1	1027	861	0.2368	-	7470A	PPB	22 Aug 2024 15:44:29	U	SMPL
P3669-03DUP - 1	1027DUP	773	0.2151	-	7470A	PPB	22 Aug 2024 15:46:47	U	SMPL
P3669-03MS - 1	1027MS	11868	2.9522	-	7470A	PPB	22 Aug 2024 15:49:04	U	SMPL
P3669-03MSD - 1	1027MSD	11180	2.7824	-	7470A	PPB	22 Aug 2024 15:51:21	U	SMPL
P3657-01 - 1	917-J-WS-081624	-107	-0.0020	-	7470A	PPB	22 Aug 2024 15:53:40	U	SMPL
PB162896BS - 1	LCSW	15488	3.8452	-	7470A	PPB	22 Aug 2024 15:55:57	U	SMPL
CCV14 - 1	CCV14	18546	4.5996	-	7470A	PPB	22 Aug 2024 15:58:12	U	SMPL
CCB14 - 1	CCB14	-105	-0.0015	-	7470A	PPB	22 Aug 2024 16:00:29	U	SMPL
PB162871TB - 1	PB162871TB	-12	0.0215	-	7470A	PPB	22 Aug 2024 16:02:47	U	SMPL
PB162873TB - 1	PB162873TB	24	0.0303	-	7470A	PPB	22 Aug 2024 16:05:02	U	SMPL
P3685-04LX5 - 1		5	0.0257	-	7470A	PPB	22 Aug 2024 16:07:17	U	SMPL
P3685-04A - 1		15245	3.7852	-	7470A	PPB	22 Aug 2024 16:09:33	U	SMPL
P3684-10LX5 - 1		-152	-0.0131	-	7470A	PPB	22 Aug 2024 16:11:48	U	SMPL
P3684-10A - 1		17589	4.3635	-	7470A	PPB	22 Aug 2024 16:14:06	U	SMPL
P3669-03LX5 - 1		-3	0.0237	-	7470A	PPB	22 Aug 2024 16:16:21	U	SMPL
P3669-03A - 1		15312	3.8018	-	7470A	PPB	22 Aug 2024 16:18:40	U	SMPL
CCV15 - 1	CCV15	19825	4.9151	-	7470A	PPB	22 Aug 2024 16:20:57	U	SMPL
CCB15 - 1	CCB15	-222	-0.0303	-	7470A	PPB	22 Aug 2024 16:23:14	U	SMPL