

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132133

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/22/2024	15:55
2	CAL2	1	Water	NA	NA	20.2	7.00	08/22/2024	15:56
3	CAL3	1	Water	NA	NA	20.3	10.02	08/22/2024	15:58
4	ICV	1	Water	NA	NA	20.2	7.01	08/22/2024	16:00
5	CCV1	1	Water	NA	NA	20.4	2.02	08/22/2024	16:20
6	P3633-02	1	Water	NA	NA	20.3	7.07	08/22/2024	16:25
7	P3718-02	1	Water	NA	NA	24.3	5.84	08/22/2024	16:30
8	P3719-01	1	Water	NA	NA	24.4	5.34	08/22/2024	16:38
9	P3719-02	1	Water	NA	NA	24.4	7.04	08/22/2024	16:40
10	P3719-03	1	Water	NA	NA	23.6	7.20	08/22/2024	16:45
11	P3719-04	1	Water	NA	NA	24.2	7.15	08/22/2024	16:50
12	P3719-05	1	Water	NA	NA	24.3	7.10	08/22/2024	16:52
13	P3719-06	1	Water	NA	NA	24.4	7.17	08/22/2024	16:59
14	P3719-07	1	Water	NA	NA	24.3	7.05	08/22/2024	17:00
15	P3719-07DUP	1	Water	NA	NA	24.4	7.06	08/22/2024	17:01
16	CCV2	1	Water	NA	NA	20.3	12.02	08/22/2024	17:05

132133

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p3718

WorkList ID : 182894

Department : Wet-Chemistry

Date : 08-22-2024 15:42:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3633-02	DSN001	Water	pH	Cool 4 deg C	PSEG04	G22	08/15/2024	9040C
P3718-02	1804	Water	pH	Cool 4 deg C	PSEG03	D21	08/22/2024	9040C
P3719-01	60359	Water	pH	Cool 4 deg C	PSEG03	D21	08/22/2024	9040C
P3719-02	60361	Water	pH	Cool 4 deg C	PSEG03	D21	08/22/2024	9040C
P3719-03	60362	Water	pH	Cool 4 deg C	PSEG03	D21	08/22/2024	9040C
P3719-04	60363	Water	pH	Cool 4 deg C	PSEG03	D21	08/22/2024	9040C
P3719-05	60364	Water	pH	Cool 4 deg C	PSEG03	D21	08/22/2024	9040C
P3719-06	60365	Water	pH	Cool 4 deg C	PSEG03	D21	08/22/2024	9040C
P3719-07	60366	Water	pH	Cool 4 deg C	PSEG03	D21	08/22/2024	9040C

132133

Date/Time 08/22/24 15:42
 Raw Sample Received by: J. C. (Son)
 Raw Sample Relinquished by: J. C. (Son)

Date/Time 08/22/24 17:30
 Raw Sample Received by: J. C. (Son)
 Raw Sample Relinquished by: J. C. (Son)