

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132145

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	08/23/2024	13:00
2	P3681-01		1	6	55.00	131.0	130.0	08/23/2024	13:30
3	P3681-02		1	4	6.00	42.8	41.8	08/23/2024	14:00
4	P3681-02DUP		1	4	6.00	42.8	41.8	08/23/2024	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB132145

WorkList Name : FP8-21

WorkList ID : 132851

Department : Wet-Chemistry

Date : 08-21-2024 09:34:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3881-01	57	Solid	Flash Point	Cool 4 deg C	PSEG03	G11	08/20/2024	1010B
P3881-02	0726-D	Solid	Flash Point	Cool 4 deg C	PSEG03	G11	08/20/2024	1010B

Date/Time 08/23/2024 12:50
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: RM cwc

Date/Time 08/23/2024 15:10
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: RM cwc