



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 8/26/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:33
In Date: 08/23/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 08/24/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132148

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3720-05	SVOC-GPC-BLANK	1	1.00	1.00	2.00	2.00	100.0	
P3720-06	PEST-GPC-BLANK	2	1.00	1.00	2.00	2.00	100.0	
P3720-07	PEST-GPC-BLANK-SPIKE	3	1.00	1.00	2.00	2.00	100.0	
P3720-08	PCB-GPC-BLANK	4	1.00	1.00	2.00	2.00	100.0	
P3720-09	PCB-GPC-BLANK-SPIKE	5	1.00	1.00	2.00	2.00	100.0	
P3720-10	SVOC-GPC2-BLANK	6	1.00	1.00	2.00	2.00	100.0	
P3720-11	PEST-GPC2-BLANK	7	1.00	1.00	2.00	2.00	100.0	
P3720-12	PEST-GPC2-BLANK-SPIKE	8	1.00	1.00	2.00	2.00	100.0	
P3720-13	PCB-GPC2-BLANK	9	1.00	1.00	2.00	2.00	100.0	
P3720-14	PCB-GPC2-BLANK-SPIKE	10	1.00	1.00	2.00	2.00	100.0	
P3725-01	HIA-705Y-1-1	11	1.00	1.00	2.00	2.00	100.0	oilc
P3725-02	HIA-705Y-1-2	12	1.00	1.00	2.00	2.00	100.0	oilc
P3725-03	HIH-244N-1-1	13	1.00	1.00	2.00	2.00	100.0	oilc
P3725-04	HIH-244N-1-2	14	1.00	1.00	2.00	2.00	100.0	oilc
P3725-05	HEH-508V-1-1	15	1.00	1.00	2.00	2.00	100.0	oilc
P3725-06	HEH-508V-1-2	16	1.00	1.00	2.00	2.00	100.0	oilc
P3726-01	MOO-SOIL-PILE	17	1.18	8.68	9.86	8.86	88.5	
P3727-01	MH-21	18	1.18	8.57	9.75	8.71	87.9	
P3727-02	MH-21-EPH	19	1.18	8.43	9.61	8.5	86.8	
P3727-03	MH-21-VOC	20	1.18	8.43	9.61	8.5	86.8	
P3728-01	SU-03-08232024	21	1.12	8.74	9.86	9.39	94.6	
P3728-02	SU-03-08232024	22	1.12	8.75	9.87	9.39	94.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$