

LB132153


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43310



Max. Conc: —

10.000 PPB

A= 0.0000e+000

B= 2.3113e-004 *slope*C= -5.9175e-003 *interceptor*

Rho= 0.9998945

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/2
0.00	0.000	0.031	0.031	161	0.000	161					=
0.05	0.050					442	0				21
0.20	0.200	0.242	0.042	1071	0.0 %	1071					-5
2.50	2.500	2.384	-0.116	10338	0.0 %	10338					1
5.00	5.000	5.029	0.029	21783	0.0 %	21783					0
7.50	7.500	7.510	0.010	32520	0.0 %	32520					0
10.0	10.000	10.004	0.004	43310	0.0 %	43310					0

LB132153

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	161	-		0 SFAM01.1	PPB	8/25/2024 8:42	Std
	0.2 S01	1071	-		0.2 SFAM01.1	PPB	8/25/2024 8:44	Std
	2.5 S02	10338	-		2.5 SFAM01.1	PPB	8/25/2024 8:46	Std
	5 S03	21783	-		5 SFAM01.1	PPB	8/25/2024 8:49	Std
	7.5 S04	32520	-		7.5 SFAM01.1	PPB	8/25/2024 8:51	Std
	10 S05	43310	-		10 SFAM01.1	PPB	8/25/2024 8:56	Std
ICV096	ICV096	17098	3.9459 -		SFAM01.1	PPB	8/25/2024 8:59	SMPL
ICB096	ICB096	-186	-0.0489 -		SFAM01.1	PPB	8/25/2024 9:01	SMPL
CCV031	CCV031	22387	5.1684 -		SFAM01.1	PPB	8/25/2024 9:03	SMPL
CCB031	CCB031	-94	-0.0276 -		SFAM01.1	PPB	8/25/2024 9:05	SMPL
PB162979BL	PBS979	260	0.0542 -		SFAM01.1	PPB	8/25/2024 9:08	SMPL
P3486-01	MEOAE1	7276	1.6758 -		SFAM01.1	PPB	8/25/2024 9:10	SMPL
P3486-02	MEOAE2	3642	0.8359 -		SFAM01.1	PPB	8/25/2024 9:12	SMPL
P3486-03	MEOAE3	11069	2.5525 -		SFAM01.1	PPB	8/25/2024 9:14	SMPL
P3486-04	MEOAD7	177	0.035 -		SFAM01.1	PPB	8/25/2024 9:17	SMPL
P3486-05	MEOAE7	6610	1.5219 -		SFAM01.1	PPB	8/25/2024 9:19	SMPL
P3486-06	MEOAE9	22653	5.2299 -		SFAM01.1	PPB	8/25/2024 9:21	SMPL
P3486-07	MEOAE4	868	0.1947 -		SFAM01.1	PPB	8/25/2024 9:24	SMPL
P3486-08	MEOAE5	8966	2.0664 -		SFAM01.1	PPB	8/25/2024 9:26	SMPL
P3486-09	MEOAE5D	8819	2.0324 -		SFAM01.1	PPB	8/25/2024 9:28	SMPL
P3486-10	MEOAE5S	21231	4.9012 -		SFAM01.1	PPB	8/25/2024 9:30	SMPL
P3486-11	MEOAE6	11538	2.6609 -		SFAM01.1	PPB	8/25/2024 9:33	SMPL
P3486-12	MEOAE8	16791	3.875 -		SFAM01.1	PPB	8/25/2024 9:35	SMPL
P3486-13	MEOAF0	2109	0.4815 -		SFAM01.1	PPB	8/25/2024 9:37	SMPL
P3486-14	MEOAF5	8406	1.937 -		SFAM01.1	PPB	8/25/2024 9:39	SMPL
P3486-16	MEOAF1	2023	0.4617 -		SFAM01.1	PPB	8/25/2024 9:42	SMPL
P3486-17	MEOAF2	650	0.1443 -		SFAM01.1	PPB	8/25/2024 9:44	SMPL
P3486-18	MEOAF3	1521	0.3456 -		SFAM01.1	PPB	8/25/2024 9:46	SMPL
P3486-19	MEOAF4	806	0.1804 -		SFAM01.1	PPB	8/25/2024 9:48	SMPL
P3486-20	MEOAF6	1315	0.298 -		SFAM01.1	PPB	8/25/2024 9:51	SMPL
CCV032	CCV032	22448	5.1825 -		SFAM01.1	PPB	8/25/2024 9:53	SMPL
CCB032	CCB032	-105	-0.0302 -		SFAM01.1	PPB	8/25/2024 9:55	SMPL
PB162980BL	PBS980	-3	-0.0066 -		SFAM01.1	PPB	8/25/2024 9:58	SMPL
P3521-01	MH1Y43	1021	0.2301 -		SFAM01.1	PPB	8/25/2024 10:00	SMPL
P3521-02	MH1Y44	2406	0.5502 -		SFAM01.1	PPB	8/25/2024 10:02	SMPL
P3521-03	MH1Y45	1113	0.2513 -		SFAM01.1	PPB	8/25/2024 10:04	SMPL
P3521-04	MH1Y45D	1081	0.2439 -		SFAM01.1	PPB	8/25/2024 10:07	SMPL
P3586-01	MDCVG3	272	0.0569 -		SFAM01.1	PPB	8/25/2024 10:11	SMPL
P3586-02	MDCVG5	31	0.0012 -		SFAM01.1	PPB	8/25/2024 10:13	SMPL
P3586-04	MDCVB5	706	0.1573 -		SFAM01.1	PPB	8/25/2024 10:16	SMPL
P3586-05	MDCVB6	589	0.1302 -		SFAM01.1	PPB	8/25/2024 10:18	SMPL
P3586-06	MDCVF0	729	0.1626 -		SFAM01.1	PPB	8/25/2024 10:20	SMPL
P3586-07	MDCVF1	634	0.1406 -		SFAM01.1	PPB	8/25/2024 10:23	SMPL
P3586-08	MDCVF2	699	0.1556 -		SFAM01.1	PPB	8/25/2024 10:25	SMPL
P3586-09	MDCVF9	161	0.0313 -		SFAM01.1	PPB	8/25/2024 10:27	SMPL
P3586-10	MDCVF9D	216	0.044 -		SFAM01.1	PPB	8/25/2024 10:29	SMPL

LB132153

INSTRUMENT ID : CV1

P3586-11	MDCVF9S	12041	2.7771 -	SFAM01.1	PPB	8/25/2024 10:32 SMPL
P3586-12	MDCVG7	76	0.0116 -	SFAM01.1	PPB	8/25/2024 10:34 SMPL
P3586-13	MDCVG9	508	0.1115 -	SFAM01.1	PPB	8/25/2024 10:36 SMPL
P3586-14	MDCVH7	557	0.1228 -	SFAM01.1	PPB	8/25/2024 10:38 SMPL
P3586-15	MDCVO3	866	0.1942 -	SFAM01.1	PPB	8/25/2024 10:41 SMPL
CCV033	CCV033	22049	5.0903 -	SFAM01.1	PPB	8/25/2024 10:43 SMPL
CCB033	CCB033	-146	-0.0397 -	SFAM01.1	PPB	8/25/2024 10:45 SMPL
P3521-05	MH1Y45S	11438	2.6378 -	SFAM01.1	PPB	8/25/2024 10:47 SMPL
CCV034	CCV034	21940	5.0651 -	SFAM01.1	PPB	8/25/2024 10:50 SMPL
CCB034	CCB034	-140	-0.0383 -	SFAM01.1	PPB	8/25/2024 10:52 SMPL