

LB132166


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43344



A= 0.0000e+000

B= 2.3247e-004 *slope*C= 1.0578e-002 *y-intercept*

Rho= 0.9994939

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/d</i>
0.00	0.000	0.020	0.020	40	0.000	40					<i>43</i>
0.05	0.050					442	0				<i>-2</i>
0.20	0.200	0.205	0.005	837	0.0 %	837					<i>3</i>
2.50	2.500	2.443	-0.057	10462	0.0 %	10462					<i>-3</i>
5.00	5.000	5.160	0.160	22150	0.0 %	22150					<i>1</i>
7.50	7.500	7.286	-0.214	31295	0.0 %	31295					
10.0	10.000	10.087	0.087	43344	0.0 %	43344					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	40	-		0 SFAM01.1	PPB	8/26/2024 10:21	Std
	0.2 S01	837	-		0.2 SFAM01.1	PPB	8/26/2024 10:23	Std
	2.5 S02	10462	-		2.5 SFAM01.1	PPB	8/26/2024 10:25	Std
	5 S03	22150	-		5 SFAM01.1	PPB	8/26/2024 10:27	Std
	7.5 S04	31295	-		7.5 SFAM01.1	PPB	8/26/2024 10:32	Std
	10 S05	43344	-		10 SFAM01.1	PPB	8/26/2024 10:35	Std
ICV097	ICV097	17063	3.9772 -		SFAM01.1	PPB	8/26/2024 10:38	SMPL
ICB097	ICB097	-135	-0.0208 -		SFAM01.1	PPB	8/26/2024 10:40	SMPL
CCV034	CCV034	22241	5.1809 -		SFAM01.1	PPB	8/26/2024 10:42	SMPL
CCB034	CCB034	-123	-0.018 -		SFAM01.1	PPB	8/26/2024 10:44	SMPL
PB162981BL	PBW981	93	0.0322 -		SFAM01.1	PPB	8/26/2024 10:50	SMPL
P3482-01	ME0AA3	1397	0.3353 -		SFAM01.1	PPB	8/26/2024 10:52	SMPL
P3482-02	ME0AA4	1361	0.327 -		SFAM01.1	PPB	8/26/2024 10:54	SMPL
P3482-03	ME0AA5	1359	0.3265 -		SFAM01.1	PPB	8/26/2024 10:57	SMPL
P3482-05	ME0AA2	1329	0.3195 -		SFAM01.1	PPB	8/26/2024 11:01	SMPL
P3482-07	ME0AB0	1464	0.3509 -		SFAM01.1	PPB	8/26/2024 11:06	SMPL
P3482-08	ME0AB7	1407	0.3377 -		SFAM01.1	PPB	8/26/2024 11:08	SMPL
P3482-09	ME0AB8	1371	0.3293 -		SFAM01.1	PPB	8/26/2024 11:10	SMPL
P3482-10	ME0AB5	1275	0.307 -		SFAM01.1	PPB	8/26/2024 11:12	SMPL
P3482-11	ME0AB6	1471	0.3525 -		SFAM01.1	PPB	8/26/2024 11:15	SMPL
P3482-13	ME0AA6	1296	0.3119 -		SFAM01.1	PPB	8/26/2024 11:19	SMPL
P3482-14	ME0AA6D	1269	0.3056 -		SFAM01.1	PPB	8/26/2024 11:21	SMPL
P3482-15	ME0AA6S	5708	1.3375 -		SFAM01.1	PPB	8/26/2024 11:24	SMPL
P3482-16	ME0AB9	1500	0.3593 -		SFAM01.1	PPB	8/26/2024 11:26	SMPL
P3482-17	ME0AA7	1364	0.3277 -		SFAM01.1	PPB	8/26/2024 11:28	SMPL
P3482-19	ME0AB3	1558	0.3728 -		SFAM01.1	PPB	8/26/2024 11:33	SMPL
CCV035	CCV035	21681	5.0508 -		SFAM01.1	PPB	8/26/2024 11:35	SMPL
CCB035	CCB035	-183	-0.032 -		SFAM01.1	PPB	8/26/2024 11:37	SMPL
P3482-20	ME0AB4	226	0.0631 -		SFAM01.1	PPB	8/26/2024 11:40	SMPL
P3482-21	ME0AA9	231	0.0643 -		SFAM01.1	PPB	8/26/2024 11:42	SMPL
P3482-22	ME0AB1	247	0.068 -		SFAM01.1	PPB	8/26/2024 11:44	SMPL
PB162982BL	PBW982	213	0.0601 -		SFAM01.1	PPB	8/26/2024 11:46	SMPL
P3522-01	MH1Y46	40	0.0199 -		SFAM01.1	PPB	8/26/2024 11:49	SMPL
P3522-02	MH1Y48	1477	0.3539 -		SFAM01.1	PPB	8/26/2024 11:51	SMPL
P3522-03	MH1Y50	469	0.1196 -		SFAM01.1	PPB	8/26/2024 11:53	SMPL
P3522-04	MH1Y52	269	0.0731 -		SFAM01.1	PPB	8/26/2024 11:55	SMPL
P3522-05	MH1Y54	173	0.0508 -		SFAM01.1	PPB	8/26/2024 11:58	SMPL
P3522-06	MH1Y56	207	0.0587 -		SFAM01.1	PPB	8/26/2024 12:00	SMPL
P3522-07	MH1Y58	41	0.0201 -		SFAM01.1	PPB	8/26/2024 12:02	SMPL
P3522-08	MH1Y60	201	0.0573 -		SFAM01.1	PPB	8/26/2024 12:04	SMPL
P3522-09	MH1Y62	-17	0.0066 -		SFAM01.1	PPB	8/26/2024 12:07	SMPL
P3522-10	MH1Y64	68	0.0264 -		SFAM01.1	PPB	8/26/2024 12:09	SMPL
P3522-11	MH1Y66	287	0.0773 -		SFAM01.1	PPB	8/26/2024 12:11	SMPL
P3522-12	MH1Y68	113	0.0368 -		SFAM01.1	PPB	8/26/2024 12:14	SMPL
P3522-13	MH1Y70	17	0.0145 -		SFAM01.1	PPB	8/26/2024 12:16	SMPL
P3522-14	MH1Y72	64	0.0255 -		SFAM01.1	PPB	8/26/2024 12:18	SMPL

P3522-15	MH1Y74	62	0.025 -	SFAM01.1	PPB	8/26/2024 12:20 SMPL
P3522-16	MH1Y76	110	0.0361 -	SFAM01.1	PPB	8/26/2024 12:23 SMPL
CCV036	CCV036	21107	4.9173 -	SFAM01.1	PPB	8/26/2024 12:25 SMPL
CCB036	CCB036	-144	-0.0229 -	SFAM01.1	PPB	8/26/2024 12:27 SMPL
P3522-17	MH1Y78	795	0.1954 -	SFAM01.1	PPB	8/26/2024 12:29 SMPL
P3522-18	MH1Y80	1010	0.2454 -	SFAM01.1	PPB	8/26/2024 12:32 SMPL
P3522-19	MH1Y80D	1007	0.2447 -	SFAM01.1	PPB	8/26/2024 12:34 SMPL
P3522-20	MH1Y80S	6779	1.5865 -	SFAM01.1	PPB	8/26/2024 12:36 SMPL
PB162983BL	PBW983	22	0.0157 -	SFAM01.1	PPB	8/26/2024 12:41 SMPL
P3523-01	MH1Y82	876	0.2142 -	SFAM01.1	PPB	8/26/2024 12:44 SMPL
P3523-02	MH1Y84	1249	0.3009 -	SFAM01.1	PPB	8/26/2024 12:46 SMPL
P3523-03	MH1Y86	912	0.2226 -	SFAM01.1	PPB	8/26/2024 12:48 SMPL
P3523-04	MH1Y88	1236	0.2979 -	SFAM01.1	PPB	8/26/2024 12:50 SMPL
P3523-05	MH1Y88D	921	0.2247 -	SFAM01.1	PPB	8/26/2024 12:53 SMPL
P3523-06	MH1Y88S	5452	1.278 -	SFAM01.1	PPB	8/26/2024 12:55 SMPL
P3523-07	MH1Y90	723	0.1787 -	SFAM01.1	PPB	8/26/2024 12:57 SMPL
P3527-01	MH1Z07	92	0.032 -	SFAM01.1	PPB	8/26/2024 12:59 SMPL
P3527-02	MH1Z08	95	0.0327 -	SFAM01.1	PPB	8/26/2024 13:02 SMPL
P3527-03	MH1Z09	86	0.0306 -	SFAM01.1	PPB	8/26/2024 13:04 SMPL
P3527-04	MH1Z10	80	0.0292 -	SFAM01.1	PPB	8/26/2024 13:06 SMPL
P3527-05	MH1Z11	118	0.038 -	SFAM01.1	PPB	8/26/2024 13:09 SMPL
P3527-06	MH1Z12	109	0.0359 -	SFAM01.1	PPB	8/26/2024 13:11 SMPL
P3527-07	MH1Z13	102	0.0343 -	SFAM01.1	PPB	8/26/2024 13:13 SMPL
P3527-08	MH1Z13D	130	0.0408 -	SFAM01.1	PPB	8/26/2024 13:15 SMPL
CCV037	CCV037	20601	4.7997 -	SFAM01.1	PPB	8/26/2024 13:18 SMPL
CCB037	CCB037	-167	-0.0282 -	SFAM01.1	PPB	8/26/2024 13:20 SMPL
P3527-09	MH1Z13S	4223	0.9923 -	SFAM01.1	PPB	8/26/2024 13:22 SMPL
P3527-10	MH1Z14	36	0.0189 -	SFAM01.1	PPB	8/26/2024 13:24 SMPL
P3527-11	MH1Z15	94	0.0324 -	SFAM01.1	PPB	8/26/2024 13:27 SMPL
P3527-12	MH1Z16	103	0.0345 -	SFAM01.1	PPB	8/26/2024 13:29 SMPL
P3527-13	MH1Z17	101	0.0341 -	SFAM01.1	PPB	8/26/2024 13:31 SMPL
P3527-14	MH1Z18	114	0.0371 -	SFAM01.1	PPB	8/26/2024 13:34 SMPL
PB162984BL	PBW984	139	0.0429 -	SFAM01.1	PPB	8/26/2024 13:36 SMPL
P3524-01	MH1Y47	140	0.0431 -	SFAM01.1	PPB	8/26/2024 13:38 SMPL
P3524-02	MH1Y49	215	0.0606 -	SFAM01.1	PPB	8/26/2024 13:40 SMPL
P3524-03	MH1Y51	136	0.0422 -	SFAM01.1	PPB	8/26/2024 13:43 SMPL
P3524-04	MH1Y53	135	0.042 -	SFAM01.1	PPB	8/26/2024 13:45 SMPL
P3524-05	MH1Y55	330	0.0873 -	SFAM01.1	PPB	8/26/2024 13:47 SMPL
P3524-06	MH1Y57	95	0.0327 -	SFAM01.1	PPB	8/26/2024 13:49 SMPL
P3524-07	MH1Y59	144	0.0441 -	SFAM01.1	PPB	8/26/2024 13:52 SMPL
P3524-08	MH1Y61	151	0.0457 -	SFAM01.1	PPB	8/26/2024 13:54 SMPL
P3524-09	MH1Y63	168	0.0496 -	SFAM01.1	PPB	8/26/2024 13:56 SMPL
P3524-10	MH1Y65	124	0.0394 -	SFAM01.1	PPB	8/26/2024 13:59 SMPL
P3524-11	MH1Y67	170	0.0501 -	SFAM01.1	PPB	8/26/2024 14:01 SMPL
P3524-12	MH1Y69	196	0.0561 -	SFAM01.1	PPB	8/26/2024 14:03 SMPL
P3524-15	MH1Y75	116	0.0375 -	SFAM01.1	PPB	8/26/2024 14:05 SMPL
CCV038	CCV038	21002	4.8929 -	SFAM01.1	PPB	8/26/2024 14:08 SMPL

LB132160

INSTRUMENT ID : CV1

CCB038	CCB038	-150	-0.0243 -	SFAM01.1	PPB	8/26/2024 14:10	SMPL
P3524-16	MH1Y77	176	0.0515 -	SFAM01.1	PPB	8/26/2024 14:12	SMPL
P3524-17	MH1Y79	191	0.055 -	SFAM01.1	PPB	8/26/2024 14:14	SMPL
P3524-18	MH1Y81	156	0.0468 -	SFAM01.1	PPB	8/26/2024 14:17	SMPL
P3524-19	MH1Y81D	154	0.0464 -	SFAM01.1	PPB	8/26/2024 14:19	SMPL
P3524-20	MH1Y81S	5633	1.3201 -	SFAM01.1	PPB	8/26/2024 14:21	SMPL
PB162985BL	PBW985	114	0.0371 -	SFAM01.1	PPB	8/26/2024 14:24	SMPL
P3525-01	MH1Y83	237	0.0657 -	SFAM01.1	PPB	8/26/2024 14:26	SMPL
P3525-02	MH1Y85	100	0.0338 -	SFAM01.1	PPB	8/26/2024 14:28	SMPL
P3525-03	MH1Y87	387	0.1005 -	SFAM01.1	PPB	8/26/2024 14:30	SMPL
P3525-04	MH1Y89	368	0.0961 -	SFAM01.1	PPB	8/26/2024 14:33	SMPL
P3525-05	MH1Y89D	189	0.0545 -	SFAM01.1	PPB	8/26/2024 14:35	SMPL
P3525-06	MH1Y89S	5393	1.2643 -	SFAM01.1	PPB	8/26/2024 14:37	SMPL
P3525-08	MH1Y92	65	0.0257 -	SFAM01.1	PPB	8/26/2024 14:39	SMPL
P3525-09	MH1Y93	141	0.0434 -	SFAM01.1	PPB	8/26/2024 14:42	SMPL
P3525-10	MH1Y94	231	0.0643 -	SFAM01.1	PPB	8/26/2024 14:44	SMPL
P3525-11	MH1Y95	184	0.0534 -	SFAM01.1	PPB	8/26/2024 14:46	SMPL
P3525-12	MH1Y96	504	0.1277 -	SFAM01.1	PPB	8/26/2024 14:49	SMPL
P3525-13	MH1Y97	226	0.0631 -	SFAM01.1	PPB	8/26/2024 14:51	SMPL
P3525-14	MH1Y98	171	0.0503 -	SFAM01.1	PPB	8/26/2024 14:53	SMPL
P3525-15	MH1Y99	162	0.0482 -	SFAM01.1	PPB	8/26/2024 14:55	SMPL
CCV039	CCV039	20894	4.8678 -	SFAM01.1	PPB	8/26/2024 14:58	SMPL
CCB039	CCB039	-115	-0.0162 -	SFAM01.1	PPB	8/26/2024 15:00	SMPL
P3525-16	MH1Z00	143	0.0438 -	SFAM01.1	PPB	8/26/2024 15:02	SMPL
P3525-17	MH1Z01	259	0.0708 -	SFAM01.1	PPB	8/26/2024 15:05	SMPL
P3525-18	MH1Z02	328	0.0868 -	SFAM01.1	PPB	8/26/2024 15:07	SMPL
P3525-19	MH1Z03	156	0.0468 -	SFAM01.1	PPB	8/26/2024 15:09	SMPL
P3525-20	MH1Z04	141	0.0434 -	SFAM01.1	PPB	8/26/2024 15:11	SMPL
P3525-21	MH1Z05	154	0.0464 -	SFAM01.1	PPB	8/26/2024 15:14	SMPL
P3525-22	MH1Z06	118	0.038 -	SFAM01.1	PPB	8/26/2024 15:16	SMPL
P3482-04		653	0.1624 -	SFAM01.1	PPB	8/26/2024 15:18	SMPL
P3482-06		565	0.1419 -	SFAM01.1	PPB	8/26/2024 15:20	SMPL
P3482-12		587	0.147 -	SFAM01.1	PPB	8/26/2024 15:23	SMPL
P3482-18		590	0.1477 -	SFAM01.1	PPB	8/26/2024 15:25	SMPL
CCV040	CCV040	21005	4.8936 -	SFAM01.1	PPB	8/26/2024 15:27	SMPL
CCB040	CCB040	-106	-0.0141 -	SFAM01.1	PPB	8/26/2024 15:30	SMPL