

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 8/27/2024

OVENTEMP IN Celsius(°C): 108
Time IN: 14:05
In Date: 08/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:38
Out Date: 08/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132163

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3733-01	A4495	1	1.14	8.64	9.78	8.52	85.4	
P3733-02	A4496	2	1.18	8.52	9.7	8.29	83.5	
P3733-03	A4497	3	1.18	8.68	9.86	8.53	84.7	
P3733-04	A4498	4	1.15	8.49	9.64	7.54	75.3	
P3733-05	A4499	5	1.15	8.83	9.98	9.09	89.9	
P3733-06	A44A0	6	1.18	8.67	9.85	8.89	88.9	
P3733-07	A44A1	7	1.18	8.43	9.61	7.29	72.5	
P3733-08	A44A2	8	1.18	8.79	9.97	8.78	86.5	
P3733-09	A44A3	9	1.15	8.83	9.98	8.77	86.3	
P3733-10	A44A3MS	10	1.15	8.83	9.98	8.77	86.3	
P3733-11	A44A3MSD	11	1.15	8.83	9.98	8.77	86.3	
P3733-12	A44A4	12	1.18	8.60	9.78	8.45	84.5	
P3733-13	A44A5	13	1.18	8.48	9.66	7.81	78.2	
P3733-14	A44A6	14	1.18	8.40	9.58	7.62	76.7	
P3733-15	A44A7	15	1.19	8.43	9.62	8.95	92.1	
P3733-16	A44A8	16	1.15	8.68	9.83	8.26	81.9	
P3733-17	A44A9	17	1.15	8.63	9.78	9.33	94.8	
P3733-18	A44B3	18	1.00	1.00	2.00	2.00	100.0	MeOH Blank
P3733-19	A44B4	19	1.00	1.00	2.00	2.00	100.0	Soil Blank
P3733-20	A44B5	20	1.15	8.82	9.97	7.34	70.2	
P3733-21	A44B7	21	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P3733-22	A44B8	22	1.00	1.00	2.00	2.00	100.0	MeOH VOA PT
P3733-23	A44B9	23	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P3733-24	VHBLK001	24	1.00	1.00	2.00	2.00	100.0	vhblk

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$