

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132165

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	08/26/2024	09:10
2	P3684-08		1	11	100.00	>212.0	>212.0	08/26/2024	10:20
3	P3684-08DUP		1	11	100.00	>212.0	>212.0	08/26/2024	10:50
4	P3684-09		1	14	100.00	>212.0	>212.0	08/26/2024	12:45
5	P3684-10		1	12	100.00	>212.0	>212.0	08/26/2024	13:15
6	P3719-01		1	12	100.00	>212.0	>212.0	08/26/2024	13:45
7	P3719-02		1	11	100.00	>212.0	>212.0	08/26/2024	14:15
8	P3719-03		1	14	100.00	>212.0	>212.0	08/26/2024	14:45
9	P3719-04		1	14	100.00	>212.0	>212.0	08/26/2024	15:15
10	P3719-05		1	14	100.00	>212.0	>212.0	08/26/2024	15:45
11	P3719-06		1	11	100.00	>212.0	>212.0	08/26/2024	16:15
12	P3719-07		1	12	100.00	>212.0	>212.0	08/26/2024	16:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132165

WorkList Name : FP-W-21 WorkList ID : 182852 Department : Wet-Chemistry Date : 08-21-2024 13:35:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3684-08	T3650	Water	Flash Point	Cool 4 deg C	PSEG03	G11	08/20/2024	1010B
P3684-09	D3643	Water	Flash Point	Cool 4 deg C	PSEG03	G11	08/20/2024	1010B
P3684-10	D3637	Water	Flash Point	Cool 4 deg C	PSEG03	G11	08/20/2024	1010B
P3719-01	60359	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/22/2024	1010B
P3719-02	60361	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/22/2024	1010B
P3719-03	60362	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/22/2024	1010B
P3719-04	60363	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/22/2024	1010B
P3719-05	60364	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/22/2024	1010B
P3719-06	60365	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/22/2024	1010B
P3719-07	60366	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/22/2024	1010B

Date/Time 08/26/2024 09:00
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: RM CWG

Date/Time 08/26/2024 16:50
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: RM CWG