

LB132175

7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42679

A= 0.0000e+000

B= 2.3859e-004

C= -7.9336e-003

Rho= 0.9991529

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/e)
0.0	0.000	0.014	0.014	94	0.000	94	0				-
0.2	0.200	0.204	0.004	888	0.0 %	888					2
2.5	2.500	2.553	0.053	10733	0.0 %	10733					+2
5.0	5.000	5.070	0.070	21281	0.0 %	21281					1
7.5	7.500	7.184	-0.316	30143	0.0 %	30143					-4
10.0	10.000	10.175	0.175	42679	0.0 %	42679					2

LB132175 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Aug 2024 12:15:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	94	-	0.0000	7470A	PPB	27 Aug 2024 12:22:15	S	Std
0.2 - 1	502	888	-	0.2000	7470A	PPB	27 Aug 2024 12:24:32	S	Std
2.5 - 1	525	10733	-	2.5000	7470A	PPB	27 Aug 2024 12:26:48	S	Std
5.0 - 1	55	21281	-	5.0000	7470A	PPB	27 Aug 2024 12:29:05	S	Std
7.5 - 1	575	30143	-	7.5000	7470A	PPB	27 Aug 2024 12:31:26	S	Std
10.0 - 1	511	42679	-	10.0000	7470A	PPB	27 Aug 2024 12:33:43	S	Std
ICV98 - 1	ICV98	16992	4.0463	-	7470A	PPB	27 Aug 2024 12:36:30	U	SMPL
ICB98 - 1	ICB98	-158	-0.0456	-	7470A	PPB	27 Aug 2024 12:38:46	U	SMPL
CCV41 - 1	CCV41	21300	5.0741	-	7470A	PPB	27 Aug 2024 12:41:03	U	SMPL
CCB41 - 1	CCB41	-162	-0.0466	-	7470A	PPB	27 Aug 2024 12:43:18	U	SMPL
CRA - 1	CRA	873	0.2004	-	7470A	PPB	27 Aug 2024 12:45:35	U	SMPL
HighStd - 1	HighStd	42805	10.2051	-	7470A	PPB	27 Aug 2024 12:47:49	U	SMPL
ChkStd - 1	ChkStd	26306	6.2685	-	7470A	PPB	27 Aug 2024 12:50:08	U	SMPL
PB163028BL - 1	PBW	0	-0.0079	-	7470A	PPB	27 Aug 2024 12:52:26	U	SMPL
ChkStd - 1	ChkStd	28360	6.7586	-	7470A	PPB	27 Aug 2024 12:52:54	U	SMPL
PB163028BL - 1	PBW	90	0.0135	-	7470A	PPB	27 Aug 2024 12:55:08	U	SMPL
PB163028BS - 1	LCSW	14762	3.5142	-	7470A	PPB	27 Aug 2024 12:57:26	U	SMPL
P3718-02 - 1	1804	3548	0.8386	-	7470A	PPB	27 Aug 2024 12:59:42	U	SMPL
P3719-01 - 1	60359	592	0.1333	-	7470A	PPB	27 Aug 2024 13:09:25	U	SMPL
P3719-01DUP - 1	60359DUP	563	0.1264	-	7470A	PPB	27 Aug 2024 13:11:41	U	SMPL
P3719-01MS - 1	60359MS	14657	3.4891	-	7470A	PPB	27 Aug 2024 13:13:58	U	SMPL
P3719-01MSD - 1	60359MSD	15161	3.6094	-	7470A	PPB	27 Aug 2024 13:16:13	U	SMPL
CCV42 - 1	CCV42	21099	5.0262	-	7470A	PPB	27 Aug 2024 13:18:29	U	SMPL
CCB42 - 1	CCB42	-118	-0.0361	-	7470A	PPB	27 Aug 2024 13:20:46	U	SMPL
P3719-01LX5 - 1		110	0.0183	-	7470A	PPB	27 Aug 2024 13:23:03	U	SMPL
P3719-01A - 1		17194	4.0945	-	7470A	PPB	27 Aug 2024 13:25:18	U	SMPL
CCV43 - 1	CCV43	21210	5.0527	-	7470A	PPB	27 Aug 2024 13:27:33	U	SMPL
CCB43 - 1	CCB43	-129	-0.0387	-	7470A	PPB	27 Aug 2024 13:29:50	U	SMPL