

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132178

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/27/2024	08:20
2	CAL2	1	Water	NA	NA	20.3	7.00	08/27/2024	08:21
3	CAL3	1	Water	NA	NA	20.3	10.02	08/27/2024	08:22
4	ICV	1	Water	NA	NA	20.3	7.02	08/27/2024	08:25
5	CCV1	1	Water	NA	NA	20.3	2.01	08/27/2024	08:30
6	P3740-01	1	Solid	20.02	20	24.3	6.59	08/27/2024	08:35
7	P3740-01DUP	1	Solid	20.03	20	24.4	6.60	08/27/2024	08:36
8	CCV2	1	Water	NA	NA	20.3	12.02	08/27/2024	08:40

WORKLIST(Hardcopy Internal Chain)

UB 132178

WorkList Name : PH P3740

WorkList ID : 182990

Department : Wet-Chemistry

Date : 08-27-2024 08:02:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3740-01	BUR-GAS-1292	Solid	pH	Cool 4 deg C	PSEG03	G12	08/26/2024	9045D

Date/Time 08/27/24 08:10

Raw Sample Received by: RM SM

Raw Sample Relinquished by: RM SM

Date/Time 08/27/24 11:00

Raw Sample Received by: RM SM

Raw Sample Relinquished by: RM SM