

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132195

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	08/28/2024	12:40
2	P3718-02		1	12	100.00	>212.0	>212.0	08/28/2024	13:10
3	P3718-02DUP		1	12	100.00	>212.0	>212.0	08/28/2024	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132195

WorkList Name : fp-8-28

WorkList ID : 183020

Department : Wet-Chemistry

Date : 08-28-2024 08:05:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3718-02	1804	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/22/2024	1010B

Date/Time 08/28/2024 12:30
Raw Sample Received by: RM CWB
Raw Sample Relinquished by: RM CWB

Date/Time 08/28/2024 14:20
Raw Sample Received by: PCOE
Raw Sample Relinquished by: RM CWB